

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011719\
 Data File : VN053447.D
 Acq On : 17 Jan 2019 20:53
 Operator : JC/SP
 Sample : PB116439ZHE#09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 28


Quant Time: Jan 18 01:09:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 03:10:53 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	931806	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1465128	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1393868	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	617049	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	479198	52.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.06%	
35) Dibromofluoromethane	7.59	113	424674	50.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.58%	
50) Toluene-d8	10.09	98	1818697	51.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.86%	
62) 4-Bromofluorobenzene	12.40	95	631346	50.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.82%	
Target Compounds						
16) Acetone	3.82	43	12999	6.016	ug/l	93
20) Methylene Chloride	4.56	84	16252	1.647	ug/l	95
43) Isopropyl Acetate	8.17	43	49276	3.570	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011719\
Data File : VN053447.D
Acq On : 17 Jan 2019 20:53
Operator : JC/SP
Sample : PB116439ZHE#09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 28

Quant Time: Jan 18 01:09:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 03:10:53 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.67 min Scan# 1834

Delta R.T. 0.00 min

Lab File: VN053447.D


Acq: 17 Jan 2019 20:53

Tot Ion:168 Resp: 931806

Ion Ratio Lower Upper




168 100

99 49.2 39.1 58.7



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO



#16

Acetone

Concen: 6.016 ug/l

RT: 3.82 min Scan# 636

Delta R.T. 0.01 min

Lab File: VN053447.D


Acq: 17 Jan 2019 20:53

Tgt Ion: 43 Resp: 12999

Ion Ratio Lower Upper



43 100


58 29.9 27.1 40.7



Abundance Ion 43.00 (42.70 to 43.70): VNO

Ion 58.00 (57.70 to 58.70): VNO





#20

Methylene Chloride

Concen: 1.647 ug/l

RT: 4.56 min Scan# 868

Delta R.T. 0.01 min

Lab File: VN053447.D

Acq: 17 Jan 2019 20:53

Tot Ion: 84 Resp: 16252


Ion Ratio Lower Upper

84 100

49 114.1 94.3 141.5

51 33.4 29.4 44.2

86 70.8 51.4 77.0






Abundance Ion 83.90 (83.60 to 84.60): VN0

Ion 48.90 (48.60 to 49.60): VN0

Ion 50.90 (50.60 to 51.60): VN0

Ion 85.90 (85.60 to 86.60): VN0



#33

1,2-Dichloroethane-d4

Concen: N.D. ug/l

RT: 8.03 min Scan# 1945

Delta R.T. 0.00 min

Lab File: VN053447.D


Acq: 17 Jan 2019 20:53

Tgt Ion: 65 Resp: 479198

Ion Ratio Lower Upper



65 100


67 54.9 0.0 111.2



Abundance Ion 64.90 (64.60 to 65.60): VN0

Ion 66.90 (66.60 to 67.60): VN0






#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 8.59 min Scan# 2119
 Delta R.T. 0.00 min
 Lab File: VN053447.D
 Acq: 17 Jan 2019 20:53

Tot Ion:114 Resp: 1465128




Ion	Ratio	Lower	Upper
114	100		
63	19.6	0.0	38.6
88	15.4	0.0	30.4



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VN


Ion 87.90 (87.60 to 88.60): VN



#35
 Dibromofluoromethane
 Concen: 50.000 ug/l
 RT: 7.59 min Scan# 1810
 Delta R.T. 0.00 min
 Lab File: VN053447.D
 Acq: 17 Jan 2019 20:53

Tgt Ion:113 Resp: 424674



Ion	Ratio	Lower	Upper
113	100		
111	102.6	81.1	121.7
192	22.9	18.2	27.2




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#43

Isopropyl Acetate

Concen: 3.570 ug/l

RT: 8.17 min Scan# 1989


Delta R.T. 0.01 min

Lab File: VN053447.D




Acq: 17 Jan 2019 20:53

Tot Ion: 43 Resp: 49276

Ion	Ratio	Lower	Upper
43	100		
61	15.8	23.9	35.9#
87	10.2	11.0	16.6#



Abundance Ion 43.00 (42.70 to 43.70): VN053399.D (-1966) (-)



#50

Toluene-d8

Concen: N.D. ug/l

RT: 10.09 min Scan# 2587


Delta R.T. 0.00 min

Lab File: VN053447.D



Acq: 17 Jan 2019 20:53


Tgt Ion: 98 Resp: 1818697

Ion	Ratio	Lower	Upper
98	100		
100	64.3	51.4	77.0



Abundance Ion 98.00 (97.70 to 98.70): VN053399.D (-2569) (-)





#62

4-Bromofluorobenzene

Concen: N.D. ug/l

RT: 12.40 min Scan# 3306

Delta R.T. 0.00 min

Lab File: VN053447.D

Acq: 17 Jan 2019 20:53


Tot Ion: 95 Resp: 631346

Ion Ratio Lower Upper

95 100

174 91.5 0.0 183.2




176 88.2 0.0 177.4



Abundance Ion 94.90 (94.60 to 95.60): VN053399.D (-3290) (-)

Ion 173.80 (173.50 to 174.50): VN053399.D (-3290) (-)

Ion 175.80 (175.50 to 176.50): VN053399.D (-3290) (-)



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.41 min Scan# 2997

Delta R.T. 0.00 min

Lab File: VN053447.D

Acq: 17 Jan 2019 20:53


Tgt Ion: 117 Resp: 1393868

Ion Ratio Lower Upper

117 100

82 52.6 42.9 64.3



119 32.5 26.0 39.0




Abundance Ion 116.90 (116.60 to 117.60): VN053399.D (-2980) (-)

Ion 82.00 (81.70 to 82.70): VN053399.D (-2980) (-)

Ion 118.90 (118.60 to 119.60): VN053399.D (-2980) (-)






#72

1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 13.35 min Scan# 3599
 Delta R.T. 0.00 min
 Lab File: VN053447.D
 Acq: 17 Jan 2019 20:53

Tot Ion	Ratio	Lower	Upper
152	100		
115	55.9	27.8	83.4
150	156.7	0.0	349.0



Abundance Ion 151.90 (151.60 to 152.60): V
 Ion 114.90 (114.60 to 115.60): V
 Ion 149.90 (149.60 to 150.60): V

