


Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011719\
Data File : VN053433.D
Acq On : 17 Jan 2019 14:17
Operator : JC/SP
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 28

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Title : SW846 8260
Last Update : Wed Jan 16 03:10:53 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.7	16938	PASS
75	95	30	60	47.5	48186	PASS
95	95	100	100	100.0	101445	PASS
96	95	5	9	6.6	6737	PASS
173	174	0.00	2	0.3	301	PASS
174	95	50	100	93.2	94536	PASS
175	174	5	9	6.9	6568	PASS
176	174	95	101	96.5	91234	PASS
177	176	5	9	6.7	6130	PASS