

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011720\
 Data File : VN059756.D
 Acq On : 17 Jan 2020 13:25
 Operator : JC/MD
 Sample : L1010-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-02-011620

Quant Time: Jan 18 01:24:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 14 08:58:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	195760	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	290029	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	256381	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	110427	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	101146	46.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.06%	
35) Dibromofluoromethane	7.57	113	88193	50.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.98%	
50) Toluene-d8	10.08	98	338036	50.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.88%	
62) 4-Bromofluorobenzene	12.40	95	115622	47.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.72%	
Target Compounds						
16) Acetone	3.79	43	15944	11.972	ug/l	98
43) Isopropyl Acetate	8.14	43	34794	8.539	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011720\
Data File : VN059756.D
Acq On : 17 Jan 2020 13:25
Operator : JC/MD
Sample : L1010-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OK-02-011620

Quant Time: Jan 18 01:24:47 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 14 08:58:21 2020
Response via : Initial Calibration

