

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011720\  
 Data File : VN059768.D  
 Acq On : 17 Jan 2020 18:06  
 Operator : JC/MD  
 Sample : L1184-02  
 Misc : 5.00mL/MSVOA N/WATER  
 ALS Vial : 21 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 MH-719-(SETTLED)

Quant Time: Jan 20 04:40:03 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N011320W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Jan 14 08:58:21 2020  
 Response via : Initial Calibration

| Internal Standards         | R.T.  | QIon | Response | Conc  | Units | Dev(Min) |
|----------------------------|-------|------|----------|-------|-------|----------|
| 1) Pentafluorobenzene      | 7.64  | 168  | 202820   | 50.00 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 8.57  | 114  | 304968   | 50.00 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 11.40 | 117  | 267189   | 50.00 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 13.34 | 152  | 112298   | 50.00 | ug/l  | 0.00     |

| System Monitoring Compounds |        |     |          |       |         |      |
|-----------------------------|--------|-----|----------|-------|---------|------|
| 33) 1,2-Dichloroethane-d4   | 8.00   | 65  | 106103   | 47.11 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 94.22%  |      |
| 35) Dibromofluoromethane    | 7.57   | 113 | 92876    | 51.07 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 102.14% |      |
| 50) Toluene-d8              | 10.08  | 98  | 353389   | 50.15 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 100.30% |      |
| 62) 4-Bromofluorobenzene    | 12.39  | 95  | 118366   | 46.60 | ug/l    | 0.00 |
| Spiked Amount               | 50.000 |     | Recovery | =     | 93.20%  |      |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011720\  
Data File : VN059768.D  
Acq On : 17 Jan 2020 18:06  
Operator : JC/MD  
Sample : L1184-02  
Misc : 5.00mL/MSVOA N/WATER  
ALS Vial : 21 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
MH-719-(SETTLED)

Quant Time: Jan 20 04:40:03 2020  
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_N\METHODS\82N011320W.M  
Quant Title : SW846 8260  
QLast Update : Tue Jan 14 08:58:21 2020  
Response via : Initial Calibration

