

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011720\
 Data File : VN059769.D
 Acq On : 17 Jan 2020 18:29
 Operator : JC/MD
 Sample : L1184-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AOC-9-NORTH-(SETTLED)

Quant Time: Jan 18 01:55:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 14 08:58:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	188948	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	283398	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	240457	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	95367	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	8.00	65	98983	47.17	ug/l	0.00
Spiked Amount			Recovery	=	94.34%	
35) Dibromofluoromethane	7.57	113	87192	51.59	ug/l	0.00
Spiked Amount			Recovery	=	103.18%	
50) Toluene-d8	10.08	98	323618	49.42	ug/l	0.00
Spiked Amount			Recovery	=	98.84%	
62) 4-Bromofluorobenzene	12.40	95	105356	44.63	ug/l	0.00
Spiked Amount			Recovery	=	89.26%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011720\
Data File : VN059769.D
Acq On : 17 Jan 2020 18:29
Operator : JC/MD
Sample : L1184-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AOC-9-NORTH-(SETTLED)

Quant Time: Jan 18 01:55:44 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011320W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 14 08:58:21 2020
Response via : Initial Calibration

