

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011722\
 Data File : VN070651.D
 Acq On : 17 Jan 2022 17:45
 Operator : JC/MD
 Sample : N1141-03MEDL2 100X
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 41862MEDL2

Quant Time: Jan 18 03:18:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1091204	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1854427	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1635640	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	563004	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	705056	44.938	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	89.880%		
35) Dibromofluoromethane	8.021	113	488543	43.541	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	87.080%		
50) Toluene-d8	10.440	98	2081765	45.418	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	90.840%		
62) 4-Bromofluorobenzene	12.731	95	744699	44.984	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	89.960%		
Target Compounds					Qvalue	
16) Acetone	4.293	43	134752	13.231	ug/l	95
31) Cyclohexane	8.083	56	59478	2.402	ug/l #	57
39) Methylcyclohexane	9.459	83	36783	1.637	ug/l #	79
40) Benzene	8.459	78	61594	1.082	ug/l	96
52) Toluene	10.505	92	726174	21.083	ug/l	97
67) Ethyl Benzene	11.843	91	149084	2.345	ug/l	95
68) m/p-Xylenes	11.950	106	187541	8.103	ug/l	94
69) o-Xylene	12.278	106	50044	2.174	ug/l	89
84) 1,2,4-Trimethylbenzene	13.369	105	15403	0.382	ug/l	96
95) Naphthalene	15.509	128	10932	2.208	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011722\
Data File : VN070651.D
Acq On : 17 Jan 2022 17:45
Operator : JC/MD
Sample : N1141-03MEDL2 100X
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
41862MEDL2

Quant Time: Jan 18 03:18:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

