

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011722\
 Data File : VN070655.D
 Acq On : 17 Jan 2022 19:27
 Operator : JC/MD
 Sample : N1099-09
 Misc : 5.77g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VACLOT-BACKFILL-TILCON-09

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/18/2022
 Supervised By : Mahesh Dadoda 01/18/2022

Quant Time: Jan 18 03:20:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	1071509	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1779720	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	1388937	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	553164	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	802754	52.106	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 104.220%			
35) Dibromofluoromethane	8.019	113	542330	50.364	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 100.720%			
50) Toluene-d8	10.440	98	2209361	50.225	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 100.460%			
62) 4-Bromofluorobenzene	12.733	95	777164	48.915	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 97.840%			
Target Compounds					Qvalue	
16) Acetone	4.317	43	29938m	2.993	ug/l	
18) Methyl Acetate	4.857	43	56057	1.938	ug/l	93
25) 2-Butanone	7.351	43	40163	3.189	ug/l	93
52) Toluene	10.505	92	7529	0.228	ug/l	96
95) Naphthalene	15.507	128	9151	2.147	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011722\
Data File : VN070655.D
Acq On : 17 Jan 2022 19:27
Operator : JC/MD
Sample : N1099-09
Misc : 5.77g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VACLOT-BACKFILL-TILCON-09

Quant Time: Jan 18 03:20:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/18/2022
Supervised By :Mahesh Dadoda 01/18/2022

