

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011722\
 Data File : VN070642.D
 Acq On : 17 Jan 2022 13:50
 Operator : JC/MD
 Sample : VN0117WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0117WBL01

Quant Time: Jan 17 17:02:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 27 18:47:12 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1142054	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1869018	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1629084	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	549811	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	724865	44.144	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	88.280%
35) Dibromofluoromethane	8.021	113	508683	44.982	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	89.960%
50) Toluene-d8	10.440	98	2119991	45.891	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	91.780%
62) 4-Bromofluorobenzene	12.731	95	740836	44.401	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.800%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011722\
Data File : VN070642.D
Acq On : 17 Jan 2022 13:50
Operator : JC/MD
Sample : VN0117WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0117WBL01

Quant Time: Jan 17 17:02:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N122721W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 27 18:47:12 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.086 min Scan# 211
Delta R.T. -0.000 min
Lab File: VN070642.D
Acq: 17 Jan 2022 13:50




Instrument :
MSVOA_N
ClientSampleId :
VN0117WBL01


Tgt Ion:168 Resp: 1142054
Ion Ratio Lower Upper
168 100
99 58.4 41.3 61.9



#33
1,2-Dichloroethane-d4
Concen: 44.144 ug/l
RT: 8.437 min Scan# 2404
Delta R.T. -0.000 min
Lab File: VN070642.D
Acq: 17 Jan 2022 13:50

Tgt Ion: 65 Resp: 724865
Ion Ratio Lower Upper
65 100
67 53.5 0.0 103.0









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.968 min Scan# 2602
Delta R.T. 0.000 min
Lab File: VN070642.D
Acq: 17 Jan 2022 13:50




Instrument :
MSVOA_N
ClientSampleId :
VN0117WBL01


Tgt Ion:114 Resp: 1869018
Ion Ratio Lower Upper
114 100
63 23.1 0.0 37.8
88 16.5 0.0 27.4



#35
Dibromofluoromethane
Concen: 44.982 ug/l
RT: 8.021 min Scan# 2249
Delta R.T. 0.000 min
Lab File: VN070642.D
Acq: 17 Jan 2022 13:50

Tgt Ion:113 Resp: 508683
Ion Ratio Lower Upper
113 100
111 103.4 82.2 123.2
192 19.8 16.9 25.3









#50
Toluene-d8
Concen: 45.891 ug/l
RT: 10.440 min Scan# 3151
Delta R.T. 0.000 min
Lab File: VN070642.D
Acq: 17 Jan 2022 13:50




Instrument :
MSVOA_N
ClientSampleId :
VN0117WBL01


Tgt Ion: 98 Resp: 2119991
Ion Ratio Lower Upper
98 100
100 64.6 50.4 75.6



#62
4-Bromofluorobenzene
Concen: 44.401 ug/l
RT: 12.731 min Scan# 4005
Delta R.T. -0.000 min
Lab File: VN070642.D
Acq: 17 Jan 2022 13:50

Tgt Ion: 95 Resp: 740836
Ion Ratio Lower Upper
95 100
174 74.8 0.0 187.2
176 71.6 0.0 181.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.744 min Scan# 3637
Delta R.T. -0.000 min
Lab File: VN070642.D
Acq: 17 Jan 2022 13:50

Instrument :
MSVOA_N
ClientSampleId :
VN0117WBL01

Tgt Ion:117 Resp: 1629084
Ion Ratio Lower Upper
117 100
82 60.1 42.4 63.6
119 32.2 25.4 38.2



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.675 min Scan# 4357
Delta R.T. 0.003 min
Lab File: VN070642.D
Acq: 17 Jan 2022 13:50

Tgt Ion:152 Resp: 549811
Ion Ratio Lower Upper
152 100
115 61.3 27.5 82.5
150 157.4 0.0 344.4

