

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011723\
 Data File : VN076275.D
 Acq On : 17 Jan 2023 13:12
 Operator : JC\MD
 Sample : PB150289TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150289TB

Quant Time: Jan 18 00:53:09 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	345935	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	546598	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	481732	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	197620	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	184448	47.619	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.240%
35) Dibromofluoromethane	8.171	113	164036	47.975	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.940%
50) Toluene-d8	10.571	98	644805	51.064	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.120%
62) 4-Bromofluorobenzene	12.853	95	199879	48.487	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.980%
Target Compounds						
					Qvalue	
16) Acetone	4.454	43	45868	36.877	ug/l	96
20) Methylene Chloride	5.289	84	23391	5.834	ug/l	95
37) Ethyl Acetate	7.565	43	33500	8.627	ug/l	96
43) Isopropyl Acetate	8.695	43	184774	29.145	ug/l #	88
95) Naphthalene	15.647	128	12878	3.047	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011723\
Data File : VN076275.D
Acq On : 17 Jan 2023 13:12
Operator : JC\MD
Sample : PB150289TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150289TB

Quant Time: Jan 18 00:53:09 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

