

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011723\
 Data File : VN076277.D
 Acq On : 17 Jan 2023 14:00
 Operator : JC\MD
 Sample : PB150289ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB150289ZHE#02

Quant Time: Jan 18 00:54:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	365579	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	588403	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	526582	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	219783	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	195208	47.689	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.380%
35) Dibromofluoromethane	8.171	113	172836	46.957	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.920%
50) Toluene-d8	10.571	98	685903	50.459	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.920%
62) 4-Bromofluorobenzene	12.853	95	219803	49.532	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.060%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	36330	27.621	ug/l	94
20) Methylene Chloride	5.283	84	19181	4.390	ug/l	97
37) Ethyl Acetate	7.571	43	34969	8.366	ug/l	96
43) Isopropyl Acetate	8.694	43	249627	36.578	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011723\
Data File : VN076277.D
Acq On : 17 Jan 2023 14:00
Operator : JC\MD
Sample : PB150289ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB150289ZHE#02

Quant Time: Jan 18 00:54:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

