

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011723\
 Data File : VN076279.D
 Acq On : 17 Jan 2023 14:48
 Operator : JC\MD
 Sample : 01159-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-E

Quant Time: Jan 18 00:55:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	352988	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	573651	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	508686	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	205212	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	190158	48.113	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.220%
35) Dibromofluoromethane	8.171	113	169507	47.237	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.480%
50) Toluene-d8	10.571	98	668894	50.473	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.940%
62) 4-Bromofluorobenzene	12.853	95	210069	48.556	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.120%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	31264	24.609	ug/l	95
20) Methylene Chloride	5.289	84	16419	3.823	ug/l	94
37) Ethyl Acetate	7.571	43	26585	6.532	ug/l	96
43) Isopropyl Acetate	8.695	43	221297	33.260	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011723\
Data File : VN076279.D
Acq On : 17 Jan 2023 14:48
Operator : JC\MD
Sample : 01159-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-E

Quant Time: Jan 18 00:55:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

