

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011725\
 Data File : VN085482.D
 Acq On : 17 Jan 2025 10:16
 Operator : JC\MD
 Sample : VN0117WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0117WBL01

Quant Time: Jan 17 22:30:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	196554	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	358389	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.859	117	314818	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	125050	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	168136	52.993	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	105.980%	
35) Dibromofluoromethane	8.159	113	128534	51.696	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	103.400%	
50) Toluene-d8	10.565	98	442327	50.071	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	100.140%	
62) 4-Bromofluorobenzene	12.847	95	142880	47.282	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	94.560%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011725\
Data File : VN085482.D
Acq On : 17 Jan 2025 10:16
Operator : JC\MD
Sample : VN0117WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0117WBL01

Quant Time: Jan 17 22:30:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

