

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011725\
 Data File : VN085488.D
 Acq On : 17 Jan 2025 12:48
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jan 17 22:33:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	180665	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	324917	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	286523	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	114971	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	156453	53.648	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 107.300%		
35) Dibromofluoromethane	8.159	113	117398	52.081	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 104.160%		
50) Toluene-d8	10.565	98	399367	49.866	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 99.740%		
62) 4-Bromofluorobenzene	12.847	95	123736	45.165	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 90.340%		

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011725\
Data File : VN085488.D
Acq On : 17 Jan 2025 12:48
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jan 17 22:33:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

