

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011725\
 Data File : VN085490.D
 Acq On : 17 Jan 2025 13:36
 Operator : JC\MD
 Sample : PB166090ZHE#01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166090ZHE#01

Quant Time: Jan 17 22:34:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	185114	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	356401	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	320476	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	122339	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	161156	53.932	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	107.860%	
35) Dibromofluoromethane	8.159	113	120249	48.634	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.260%	
50) Toluene-d8	10.565	98	448488	51.052	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.100%	
62) 4-Bromofluorobenzene	12.847	95	135555	45.108	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	90.220%	
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	47767	49.474	ug/l	100
20) Methylene Chloride	5.283	84	3793	1.587	ug/l #	72
30) Chloroform	7.965	83	48226	10.695	ug/l	93
43) Isopropyl Acetate	8.683	43	160325	28.566	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011725\
Data File : VN085490.D
Acq On : 17 Jan 2025 13:36
Operator : JC\MD
Sample : PB166090ZHE#01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166090ZHE#01

Quant Time: Jan 17 22:34:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

