

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011725\
 Data File : VN085492.D
 Acq On : 17 Jan 2025 14:24
 Operator : JC\MD
 Sample : PB166090ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166090ZHE#03

Quant Time: Jan 17 22:35:37 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	187273	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	365406	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	324540	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	121820	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	166868	55.200	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	110.400%	
35) Dibromofluoromethane	8.165	113	121903	48.088	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	96.180%	
50) Toluene-d8	10.565	98	452140	50.199	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	100.400%	
62) 4-Bromofluorobenzene	12.847	95	140205	45.506	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	91.020%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	116318	119.087	ug/l	97
20) Methylene Chloride	5.277	84	2450	1.013	ug/l	95
25) 2-Butanone	7.482	43	11282	7.849	ug/l	96
43) Isopropyl Acetate	8.682	43	222983	38.751	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011725\
Data File : VN085492.D
Acq On : 17 Jan 2025 14:24
Operator : JC\MD
Sample : PB166090ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166090ZHE#03

Quant Time: Jan 17 22:35:37 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

