

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011725\
 Data File : VN085494.D
 Acq On : 17 Jan 2025 15:12
 Operator : JC\MD
 Sample : PB166090ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB166090ZHE#05

Quant Time: Jan 17 22:36:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	178617	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	353077	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	312002	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	118223	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	159606	55.356	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	110.720%	
35) Dibromofluoromethane	8.165	113	117406	47.931	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	95.860%	
50) Toluene-d8	10.565	98	438055	50.334	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	100.660%	
62) 4-Bromofluorobenzene	12.847	95	132123	44.380	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	88.760%	
Target Compounds						
					Qvalue	
16) Acetone	4.424	43	123044	132.078	ug/l	100
18) Methyl Acetate	5.036	43	3665	1.294	ug/l #	85
20) Methylene Chloride	5.283	84	2462	1.068	ug/l #	76
25) 2-Butanone	7.488	43	12704	9.267	ug/l	95
43) Isopropyl Acetate	8.688	43	268427	48.277	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011725\
Data File : VN085494.D
Acq On : 17 Jan 2025 15:12
Operator : JC\MD
Sample : PB166090ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB166090ZHE#05

Quant Time: Jan 17 22:36:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

