

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011725\
 Data File : VN085498.D
 Acq On : 17 Jan 2025 16:48
 Operator : JC\MD
 Sample : Q1115-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VNJ-213

Quant Time: Jan 17 22:38:33 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	178354	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	357752	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	323798	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	122676	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	161451	56.079	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 112.160%			
35) Dibromofluoromethane	8.165	113	119637	48.203	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 96.400%			
50) Toluene-d8	10.565	98	441325	50.047	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 100.100%			
62) 4-Bromofluorobenzene	12.847	95	136447	45.234	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 90.460%			
Target Compounds						
16) Acetone	4.430	43	126672	136.173	ug/l	95
18) Methyl Acetate	5.030	43	3367	1.190	ug/l #	72
20) Methylene Chloride	5.265	84	2444	1.061	ug/l #	79
25) 2-Butanone	7.489	43	12717	9.290	ug/l	96
43) Isopropyl Acetate	8.683	43	203410	36.106	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011725\
Data File : VN085498.D
Acq On : 17 Jan 2025 16:48
Operator : JC\MD
Sample : Q1115-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VNJ-213

Quant Time: Jan 17 22:38:33 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

