

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011725\
 Data File : VN085502.D
 Acq On : 17 Jan 2025 18:24
 Operator : JC\MD
 Sample : Q1124-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-WATER-REF-5

Quant Time: Jan 17 22:40:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 15 02:16:08 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	180360	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	353519	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	291193	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	121733	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	161818	55.581	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	111.160%	
35) Dibromofluoromethane	8.165	113	120836	49.269	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.540%	
50) Toluene-d8	10.565	98	412604	47.350	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	94.700%	
62) 4-Bromofluorobenzene	12.847	95	144949	48.628	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	97.260%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011725\
Data File : VN085502.D
Acq On : 17 Jan 2025 18:24
Operator : JC\MD
Sample : Q1124-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-WATER-REF-5

Quant Time: Jan 17 22:40:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011425W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 15 02:16:08 2025
Response via : Initial Calibration

