

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011819\
 Data File : VN053460.D
 Acq On : 18 Jan 2019 9:29
 Operator : JC/SP
 Sample : VN0118WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0118WBL01

Quant Time: Jan 18 09:56:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N122818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Dec 28 10:33:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	127288	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	690227	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	609931	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.03	65	242973	30.05	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	100.17%
60) 4-Bromofluorobenzene	12.40	95	258281	27.67	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	92.23%
63) Toluene-d8	10.09	98	810740	29.50	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	98.33%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011819\
Data File : VN053460.D
Acq On : 18 Jan 2019 9:29
Operator : JC/SP
Sample : VN0118WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
VN0118WBL01

Quant Time: Jan 18 09:56:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N122818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Dec 28 10:33:50 2018
Response via : Initial Calibration

