

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011819\
 Data File : VN053462.D
 Acq On : 18 Jan 2019 10:27
 Operator : JC/SP
 Sample : K1166-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 D-10

Quant Time: Jan 18 10:59:05 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N122818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Dec 28 10:33:50 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	103089	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	542282	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	483401	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	198589	30.32	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	101.07%
60) 4-Bromofluorobenzene	12.40	95	200790	27.15	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	90.50%
63) Toluene-d8	10.09	98	632355	29.04	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	96.80%

Target Compounds					Ovalue
15) Acetone	3.82	58	5897	13.248 ug/l	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011819\
Data File : VN053462.D
Acq On : 18 Jan 2019 10:27
Operator : JC/SP
Sample : K1166-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
D-10

Quant Time: Jan 18 10:59:05 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N122818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Dec 28 10:33:50 2018
Response via : Initial Calibration

