

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011819\
 Data File : VN053471.D
 Acq On : 18 Jan 2019 14:50
 Operator : JC/SP
 Sample : K1199-18
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 EB03-20190116

Quant Time: Jan 21 10:52:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 03:10:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	885579	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1383372	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1294723	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	563803	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	451337	51.56	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.12%	
35) Dibromofluoromethane	7.59	113	403199	51.07	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.14%	
50) Toluene-d8	10.09	98	1714210	51.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.68%	
62) 4-Bromofluorobenzene	12.40	95	583358	49.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.64%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011819\
Data File : VN053471.D
Acq On : 18 Jan 2019 14:50
Operator : JC/SP
Sample : K1199-18
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
EB03-20190116

Quant Time: Jan 21 10:52:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 03:10:53 2019
Response via : Initial Calibration

