

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011819\
 Data File : VN053472.D
 Acq On : 18 Jan 2019 15:17
 Operator : JC/SP
 Sample : K1199-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-TW521-3337

Quant Time: Jan 19 00:55:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 03:10:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	904345	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1411925	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1295894	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	529089	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	455710	50.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.96%	
35) Dibromofluoromethane	7.59	113	418804	51.97	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.94%	
50) Toluene-d8	10.09	98	1736710	50.96	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.92%	
62) 4-Bromofluorobenzene	12.40	95	571701	47.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.68%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011819\
Data File : VN053472.D
Acq On : 18 Jan 2019 15:17
Operator : JC/SP
Sample : K1199-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SA-TW521-3337

Quant Time: Jan 19 00:55:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 03:10:53 2019
Response via : Initial Calibration

