

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011819\
 Data File : VN053479.D
 Acq On : 18 Jan 2019 18:24
 Operator : JC/SP
 Sample : K1199-17
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SA-TW529-3337

Quant Time: Jan 21 10:54:53 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 03:10:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	837822	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1314323	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1231825	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	543211	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	431777	52.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.28%	
35) Dibromofluoromethane	7.59	113	395490	52.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.44%	
50) Toluene-d8	10.09	98	1619261	51.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.08%	
62) 4-Bromofluorobenzene	12.40	95	550893	49.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.04%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011819\
Data File : VN053479.D
Acq On : 18 Jan 2019 18:24
Operator : JC/SP
Sample : K1199-17
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SA-TW529-3337

Quant Time: Jan 21 10:54:53 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 03:10:53 2019
Response via : Initial Calibration

