

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011819\
 Data File : VN053491.D
 Acq On : 18 Jan 2019 23:45
 Operator : JC/SP
 Sample : K1011-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 34 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-03-011719-A

Quant Time: Jan 19 02:08:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jan 16 03:10:53 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	827980	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1310800	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1271705	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	572083	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	433748	53.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.00%	
35) Dibromofluoromethane	7.59	113	388346	51.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.82%	
50) Toluene-d8	10.09	98	1647293	52.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.14%	
62) 4-Bromofluorobenzene	12.40	95	579764	52.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.52%	
Target Compounds						
16) Acetone	3.82	43	12370	6.442	ug/l	96
20) Methylene Chloride	4.56	84	590905	67.408	ug/l	99
43) Isopropyl Acetate	8.17	43	44130	3.573	ug/l #	81
95) Naphthalene	15.13	128	19929	5.266	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN011819\
Data File : VN053491.D
Acq On : 18 Jan 2019 23:45
Operator : JC/SP
Sample : K1011-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 34 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
OR-03-011719-A

Quant Time: Jan 19 02:08:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
Quant Title : SW846 8260
QLast Update : Wed Jan 16 03:10:53 2019
Response via : Initial Calibration

