


Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN011819\
 Data File : VN053458.D
 Acq On : 18 Jan 2019 8:21
 Operator : JC/SP
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N011519W.M
 Title : SW846 8260
 Last Update : Wed Jan 16 03:10:53 2019



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.3	12560	PASS
75	95	30	60	46.1	35557	PASS
95	95	100	100	100.0	77104	PASS
96	95	5	9	6.5	5007	PASS
173	174	0.00	2	1.1	781	PASS
174	95	50	100	92.9	71632	PASS
175	174	5	9	7.6	5447	PASS
176	174	95	101	95.0	68064	PASS
177	176	5	9	6.8	4620	PASS