

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011822\
 Data File : VN070668.D
 Acq On : 18 Jan 2022 13:51
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Quant Time: Jan 19 01:10:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	110	0.00
2 T	Dichlorodifluoromethane	0.578	0.549	5.0	108	0.00
3 P	Chloromethane	0.845	0.776	8.2	106	0.00
4 C	Vinyl Chloride	0.803	0.724	9.8#	101	0.00
5 T	Bromomethane	0.418	0.402	3.8	107	0.00
6 T	Chloroethane	0.501	0.416	17.0	96	0.00
7 T	Trichlorofluoromethane	0.918	0.864	5.9	107	0.00
8 T	Diethyl Ether	0.401	0.396	1.2	113	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.555	0.516	7.0	108	0.00
10 T	Methyl Iodide	0.715	0.684	4.3	103	0.00
11 T	Tert butyl alcohol	0.129	0.120	7.0	106	0.00
12 CM	1,1-Dichloroethene	0.553	0.512	7.4#	108	0.00
13 T	Acrolein	0.185	0.147	20.5	95	0.00
14 T	Allyl chloride	1.205	1.192	1.1	107	0.00
15 T	Acrylonitrile	0.411	0.394	4.1	103	0.00
16 T	Acetone	0.497	0.437	12.1	96	0.00
17 T	Carbon Disulfide	1.502	1.454	3.2	107	0.00
18 T	Methyl Acetate	0.978	0.896	8.4	104	0.00
19 T	Methyl tert-butyl Ether	1.976	1.941	1.8	106	0.00
20 T	Methylene Chloride	0.655	0.613	6.4	106	0.00
21 T	trans-1,2-Dichloroethene	0.557	0.542	2.7	106	0.00
22 T	Diisopropyl ether	2.389	2.348	1.7	107	0.00
23 T	Vinyl Acetate	2.025	2.042	-0.8	107	0.00
24 P	1,1-Dichloroethane	1.195	1.153	3.5	106	0.00
25 T	2-Butanone	0.626	0.593	5.3	101	0.00
26 T	2,2-Dichloropropane	0.916	0.905	1.2	109	0.00
27 T	cis-1,2-Dichloroethene	0.661	0.641	3.0	107	0.00
28 T	Bromochloromethane	0.536	0.499	6.9	106	0.00
29 T	Tetrahydrofuran	0.383	0.367	4.2	102	0.00
30 C	Chloroform	1.342	1.093	18.6#	106	0.00
31 T	Cyclohexane	1.262	1.186	6.0	108	0.00
32 T	1,1,1-Trichloroethane	0.927	0.906	2.3	107	0.00
33 S	1,2-Dichloroethane-d4	0.717	0.673	6.1	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	109	0.00
35 S	Dibromofluoromethane	0.310	0.302	2.6	105	0.00
36 T	1,1-Dichloropropene	0.522	0.523	-0.2	109	0.00
37 T	Ethyl Acetate	0.693	0.661	4.6	101	0.00
38 T	Carbon Tetrachloride	0.464	0.462	0.4	107	0.00
39 T	Methylcyclohexane	0.645	0.673	-4.3	110	0.00
40 TM	Benzene	1.618	1.579	2.4	106	0.00
41 T	Methacrylonitrile	0.353	0.352	0.3	109	0.00
42 TM	1,2-Dichloroethane	0.561	0.544	3.0	106	0.00
43 T	Isopropyl Acetate	1.059	1.059	0.0	107	0.00
44 TM	Trichloroethene	0.367	0.363	1.1	107	0.00
45 C	1,2-Dichloropropane	0.440	0.433	1.6#	107	0.00
46 T	Dibromomethane	0.256	0.254	0.8	106	0.00
47 T	Bromodichloromethane	0.523	0.510	2.5	106	0.00
48 T	Methyl methacrylate	0.504	0.526	-4.4	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011822\
 Data File : VN070668.D
 Acq On : 18 Jan 2022 13:51
 Operator : JC/MD
 Sample : VSTDICV050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Quant Time: Jan 19 01:10:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	106	0.00
50 S	Toluene-d8	1.307	1.266	3.1	105	0.00
51 T	4-Methyl-2-Pentanone	0.676	0.674	0.3	104	0.00
52 CM	Toluene	0.968	0.969	-0.1#	107	0.00
53 T	t-1,3-Dichloropropene	0.561	0.590	-5.2	109	0.00
54 T	cis-1,3-Dichloropropene	0.627	0.650	-3.7	108	0.00
55 T	1,1,2-Trichloroethane	0.380	0.367	3.4	105	0.00
56 T	Ethyl methacrylate	0.639	0.661	-3.4	106	0.00
57 T	1,3-Dichloropropane	0.673	0.666	1.0	106	0.00
58 T	2-Chloroethyl Vinyl ether	0.151	0.148	2.0	97	0.00
59 T	2-Hexanone	0.491	0.504	-2.6	103	0.00
60 T	Dibromochloromethane	0.355	0.363	-2.3	106	0.00
61 T	1,2-Dibromoethane	0.382	0.376	1.6	105	0.00
62 S	4-Bromofluorobenzene	0.466	0.476	-2.1	109	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	110	0.00
64 T	Tetrachloroethene	0.367	0.349	4.9	106	0.00
65 PM	Chlorobenzene	1.083	1.051	3.0	107	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.366	1.9	106	0.00
67 C	Ethyl Benzene	2.008	2.017	-0.4#	108	0.00
68 T	m/p-Xylenes	0.731	0.744	-1.8	108	0.00
69 T	o-Xylene	0.725	0.732	-1.0	107	0.00
70 T	Styrene	1.160	1.212	-4.5	107	0.00
71 P	Bromoform	0.282	0.288	-2.1	107	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
73 T	Isopropylbenzene	4.730	4.444	6.0	108	0.00
74 T	N-amyl acetate	1.934	1.933	0.1	106	0.00
75 P	1,1,2,2-Tetrachloroethane	1.598	1.418	11.3	103	0.00
76 T	1,2,3-Trichloropropane	1.326	1.247	6.0	113	0.00
77 T	Bromobenzene	1.061	0.999	5.8	107	0.00
78 T	n-propylbenzene	5.290	5.227	1.2	109	0.00
79 T	2-Chlorotoluene	3.346	3.172	5.2	108	0.00
80 T	1,3,5-Trimethylbenzene	3.825	3.699	3.3	108	0.00
81 T	trans-1,4-Dichloro-2-butene	0.445	0.446	-0.2	109	0.00
82 T	4-Chlorotoluene	3.127	3.090	1.2	109	0.00
83 T	tert-Butylbenzene	3.330	3.204	3.8	108	0.00
84 T	1,2,4-Trimethylbenzene	3.677	3.626	1.4	108	0.00
85 T	sec-Butylbenzene	4.527	4.527	0.0	109	0.00
86 T	p-Isopropyltoluene	3.556	3.651	-2.7	109	0.00
87 T	1,3-Dichlorobenzene	1.764	1.737	1.5	108	0.00
88 T	1,4-Dichlorobenzene	1.756	1.681	4.3	108	0.00
89 T	n-Butylbenzene	2.934	3.098	-5.6	112	0.00
90 T	Hexachloroethane	0.622	0.613	1.4	108	0.00
91 T	1,2-Dichlorobenzene	1.743	1.645	5.6	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.271	0.250	7.7	103	0.00
93 T	1,2,4-Trichlorobenzene	0.774	0.865	-11.8	112	0.00
94 T	Hexachlorobutadiene	0.516	0.520	-0.8	111	0.00
95 T	Naphthalene	2.001	2.361	-18.0	114	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011822\
Data File : VN070668.D
Acq On : 18 Jan 2022 13:51
Operator : JC/MD
Sample : VSTDICV050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Quant Time: Jan 19 01:10:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.738	0.832	-12.7	111	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6