

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011822\
 Data File : VN070665.D
 Acq On : 18 Jan 2022 12:34
 Operator : JC/MD
 Sample : VSTDIC100
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jan 18 12:53:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 12:52:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.083	168	844621	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1398268	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.741	117	1322861	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	567560	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	1258809	103.749	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 207.500%#	
35) Dibromofluoromethane	8.021	113	915436	107.635	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 215.260%#	
50) Toluene-d8	10.440	98	3830964	109.462	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 218.920%#	
62) 4-Bromofluorobenzene	12.728	95	1452643	115.397	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 230.800%#	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.073	85	993342	104.797	ug/l	100
3) Chloromethane	2.303	50	1424423	108.109	ug/l	99
4) Vinyl Chloride	2.448	62	1388519	114.582	ug/l	99
5) Bromomethane	2.832	94	723185	116.661	ug/l	99
6) Chloroethane	3.006	64	833170	114.038	ug/l	98
7) Trichlorofluoromethane	3.371	101	1566581	108.792	ug/l	98
8) Diethyl Ether	3.821	74	742563	117.260	ug/l	82
9) 1,1,2-Trichlorotrifluo...	4.205	101	1018941	116.757	ug/l	96
10) Methyl Iodide	4.419	142	1325426	112.673	ug/l	98
11) Tert butyl alcohol	5.379	59	1179378	578.534	ug/l	98
12) 1,1-Dichloroethene	4.178	96	1033090	119.964	ug/l	93
13) Acrolein	4.038	56	1734325	913.555	ug/l	100
14) Allyl chloride	4.838	41	2233266	113.520	ug/l #	83
15) Acrylonitrile	5.551	53	3726800	580.517	ug/l	99
16) Acetone	4.283	43	4563563	577.334	ug/l	97
17) Carbon Disulfide	4.527	76	2706737	114.204	ug/l	98
18) Methyl Acetate	4.851	43	1679509	77.439	ug/l	92
19) Methyl tert-butyl Ether	5.615	73	3560672	110.916	ug/l	94
20) Methylene Chloride	5.092	84	1129277	103.276	ug/l	87
21) trans-1,2-Dichloroethene	5.594	96	995164	108.879	ug/l	86
22) Diisopropyl ether	6.495	45	4249387	110.708	ug/l	91
23) Vinyl Acetate	6.431	43	18737374	619.743	ug/l	95
24) 1,1-Dichloroethane	6.385	63	2139489	108.239	ug/l	98
25) 2-Butanone	7.332	43	5626844	561.287	ug/l	92
26) 2,2-Dichloropropane	7.324	77	1662663	107.367	ug/l	97
27) cis-1,2-Dichloroethene	7.324	96	1181956	108.724	ug/l	86
28) Bromochloromethane	7.656	49	963918	101.668	ug/l #	75
29) Tetrahydrofuran	7.681	42	3449369	574.009	ug/l	87
30) Chloroform	7.815	83	2007351	105.635	ug/l	99
31) Cyclohexane	8.094	56	2158017	111.097	ug/l	86
32) 1,1,1-Trichloroethane	8.013	97	1687108	105.992	ug/l	97
36) 1,1-Dichloropropene	8.220	75	1559139	110.854	ug/l	96
37) Ethyl Acetate	7.410	43	2045199	111.315	ug/l	98
38) Carbon Tetrachloride	8.206	117	1411893	106.222	ug/l	100
39) Methylcyclohexane	9.459	83	1994381	116.466	ug/l	89
40) Benzene	8.458	78	4718645	109.174	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011822\
 Data File : VN070665.D
 Acq On : 18 Jan 2022 12:34
 Operator : JC/MD
 Sample : VSTDIC100
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Quant Time: Jan 18 12:53:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 12:52:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.635	41	1094701	120.014	ug/l #	85
42) 1,2-Dichloroethane	8.528	62	1624689	101.130	ug/l	99
43) Isopropyl Acetate	8.560	43	3245677	112.621	ug/l	96
44) Trichloroethene	9.215	130	1090608	106.696	ug/l	89
45) 1,2-Dichloropropane	9.488	63	1300956	111.608	ug/l	100
46) Dibromomethane	9.574	93	761047	105.886	ug/l	94
47) Bromodichloromethane	9.759	83	1569285	110.190	ug/l	99
48) Methyl methacrylate	9.558	41	1609276	117.580	ug/l #	78
49) 1,4-Dioxane	9.569	88	467797	2304.518	ug/l #	83
51) 4-Methyl-2-Pentanone	10.330	43	10404798	573.497	ug/l	99
52) Toluene	10.505	92	2902253	111.178	ug/l	97
53) t-1,3-Dichloropropene	10.717	75	1825800	118.006	ug/l	99
54) cis-1,3-Dichloropropene	10.188	75	1992424	118.535	ug/l	99
55) 1,1,2-Trichloroethane	10.896	97	1117732	108.358	ug/l	97
56) Ethyl methacrylate	10.759	69	2087410	121.748	ug/l	88
57) 1,3-Dichloropropane	11.041	76	2022131	110.524	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.041	63	2634912	717.593	ug/l	90
59) 2-Hexanone	11.084	43	7804442	589.035	ug/l	97
60) Dibromochloromethane	11.237	129	1136352	113.711	ug/l	100
61) 1,2-Dibromoethane	11.344	107	1169636	111.747	ug/l	99
64) Tetrachloroethene	10.977	164	957724	99.656	ug/l	94
65) Chlorobenzene	11.768	112	3037371	108.091	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.840	131	1060539	106.014	ug/l	99
67) Ethyl Benzene	11.843	91	5737805	111.114	ug/l	96
68) m/p-Xylenes	11.950	106	4200500	223.309	ug/l	92
69) o-Xylene	12.277	106	2102107	112.530	ug/l	91
70) Styrene	12.291	104	3526920	119.059	ug/l	96
71) Bromoform	12.454	173	867833	117.071	ug/l #	99
73) Isopropylbenzene	12.575	105	5523223	105.055	ug/l	98
74) N-amyl acetate	12.385	43	2470801	111.122	ug/l	95
75) 1,1,2,2-Tetrachloroethane	12.825	83	1730408	97.541	ug/l	99
76) 1,2,3-Trichloropropane	12.878	75	2231718	132.790	ug/l #	1
77) Bromobenzene	12.859	156	1248978	103.688	ug/l	75
78) n-propylbenzene	12.918	91	6517578	111.369	ug/l	96
79) 2-Chlorotoluene	13.004	91	3917481	104.575	ug/l	92
80) 1,3,5-Trimethylbenzene	13.058	105	4586276	108.668	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.623	75	581194	124.540	ug/l	90
82) 4-Chlorotoluene	13.101	91	3835336	108.955	ug/l	91
83) tert-Butylbenzene	13.321	119	4011118	107.867	ug/l	94
84) 1,2,4-Trimethylbenzene	13.366	105	4480247	110.185	ug/l	97
85) sec-Butylbenzene	13.500	105	5671976	112.240	ug/l	97
86) p-Isopropyltoluene	13.613	119	4539494	114.277	ug/l	96
87) 1,3-Dichlorobenzene	13.613	146	2112428	104.249	ug/l	96
88) 1,4-Dichlorobenzene	13.694	146	2068111	103.375	ug/l	97
89) n-Butylbenzene	13.940	91	3921354	118.223	ug/l	97
90) Hexachloroethane	14.209	117	794802	115.702	ug/l	92
91) 1,2-Dichlorobenzene	13.986	146	2009037	101.184	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	14.600	75	332206	108.822	ug/l	82
93) 1,2,4-Trichlorobenzene	15.263	180	1083602	115.003	ug/l	98
94) Hexachlorobutadiene	15.367	225	586601	106.239	ug/l	99
95) Naphthalene	15.504	128	3063778	119.941	ug/l	100
96) 1,2,3-Trichlorobenzene	15.697	180	1036364	112.661	ug/l	99


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011822\
Data File : VN070665.D
Acq On : 18 Jan 2022 12:34
Operator : JC/MD
Sample : VSTDICC100
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1


Quant Time: Jan 18 12:53:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 12:52:54 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed





Quant Time: Jan 18 12:53:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 12:52:54 2022
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.083 min Scan# 2272
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion: 168 Resp: 844621
Ion Ratio Lower Upper
168 100
99 58.4 41.3 61.9



#2
Dichlorodifluoromethane
Concen: 104.797 ug/l
RT: 2.073 min Scan# 31
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion: 85 Resp: 993342
Ion Ratio Lower Upper
85 100
87 32.8 16.3 48.9








#3
Chloromethane
Concen: 108.109 ug/l
RT: 2.303 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion: 50 Resp: 1424423
Ion Ratio Lower Upper
50 100
52 32.1 25.4 38.0



#4
Vinyl Chloride
Concen: 114.582 ug/l
RT: 2.448 min Scan# 171
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion: 62 Resp: 1388519
Ion Ratio Lower Upper
62 100
64 31.5 25.8 38.6








#5
Bromomethane
Concen: 116.661 ug/l
RT: 2.832 min Scan# 314
Delta R.T. -0.016 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion: 94 Resp: 723185
Ion Ratio Lower Upper
94 100
96 92.9 73.8 110.8



#6
Chloroethane
Concen: 114.038 ug/l
RT: 3.006 min Scan# 379
Delta R.T. -0.014 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion: 64 Resp: 833170
Ion Ratio Lower Upper
64 100
66 31.8 26.3 39.5








#7
 Trichlorofluoromethane
 Concen: 108.792 ug/l
 RT: 3.371 min Scan# 515
 Delta R.T. -0.006 min
 Lab File: VN070665.D
 Acq: 18 Jan 2022 12:34


Tgt Ion:101 Resp: 1566581
 Ion Ratio Lower Upper
 101 100
 103 63.9 50.1 75.1



#8
 Diethyl Ether
 Concen: 117.260 ug/l
 RT: 3.821 min Scan# 683
 Delta R.T. -0.003 min
 Lab File: VN070665.D
 Acq: 18 Jan 2022 12:34

Tgt Ion: 74 Resp: 742563
 Ion Ratio Lower Upper
 74 100
 45 111.3 47.0 141.2









#9
1,1,2-Trichlorotrifluoroethane
Concen: 116.757 ug/l
RT: 4.205 min Scan# 826
Delta R.T. -0.006 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion:101 Resp: 1018941




Ion	Ratio	Lower	Upper
101	100		
85	44.7	37.2	55.8
151	77.8	65.9	98.9




#10
Methyl Iodide
Concen: 112.673 ug/l
RT: 4.419 min Scan# 906
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion:142 Resp: 1325426





Ion	Ratio	Lower	Upper
142	100		
127	43.0	35.4	53.0
141	14.3	11.6	17.4








#11
Tert butyl alcohol
Concen: 578.534 ug/l
RT: 5.379 min Scan# 1264
Delta R.T. 0.003 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion: 59 Resp: 1179378
Ion Ratio Lower Upper
59 100
57 10.0 8.7 13.1



#12
1,1-Dichloroethene
Concen: 119.964 ug/l
RT: 4.178 min Scan# 816
Delta R.T. -0.003 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion: 96 Resp: 1033090
Ion Ratio Lower Upper
96 100
61 175.0 131.0 196.6
98 63.8 52.4 78.6








#13
Acrolein
Concen: 913.555 ug/l
RT: 4.038 min Scan# 764
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion: 56 Resp: 1734325
Ion Ratio Lower Upper
56 100
55 70.7 56.6 85.0



#14
Allyl chloride
Concen: 113.520 ug/l
RT: 4.838 min Scan# 1062
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 41 Resp: 2233266
Ion Ratio Lower Upper
41 100
39 59.7 61.2 91.8#
76 27.2 26.0 39.0









#15
Acrylonitrile
Concen: 580.517 ug/l
RT: 5.551 min Scan# 1328
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 53 Resp: 3726800




Ion	Ratio	Lower	Upper
53	100		
52	82.4	66.2	99.2
51	35.2	29.6	44.4




#16
Acetone
Concen: 577.334 ug/l
RT: 4.283 min Scan# 855
Delta R.T. -0.003 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 43 Resp: 4563563





Ion	Ratio	Lower	Upper
43	100		
58	30.6	23.1	34.7








#17
Carbon Disulfide
Concen: 114.204 ug/l
RT: 4.527 min Scan# 946
Delta R.T. -0.003 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion: 76 Resp: 2706737
Ion Ratio Lower Upper
76 100
78 8.8 7.5 11.3



#18
Methyl Acetate
Concen: 77.439 ug/l
RT: 4.851 min Scan# 1067
Delta R.T. -0.003 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 43 Resp: 1679509
Ion Ratio Lower Upper
43 100
74 20.1 19.4 29.2









#19

Methyl tert-butyl Ether
Concen: 110.916 ug/l
RT: 5.615 min Scan# 1352
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34




Tgt Ion: 73 Resp: 3560672
Ion Ratio Lower Upper
73 100
57 23.8 16.8 25.2



#20

Methylene Chloride
Concen: 103.276 ug/l
RT: 5.092 min Scan# 1157
Delta R.T. -0.003 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 84 Resp: 1129277
Ion Ratio Lower Upper
84 100
49 139.7 93.5 140.3
51 42.3 29.4 44.2
86 65.0 53.0 79.4









#21
trans-1,2-Dichloroethene
Concen: 108.879 ug/l
RT: 5.594 min Scan# 1344
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 96 Resp: 995164




Ion	Ratio	Lower	Upper
96	100		
61	155.6	105.8	158.6
98	62.5	52.0	78.0




#22
Diisopropyl ether
Concen: 110.708 ug/l
RT: 6.495 min Scan# 1680
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 45 Resp: 4249387

Ion	Ratio	Lower	Upper
45	100		
43	57.1	51.7	77.5
87	23.4	23.2	34.8
59	10.8	9.4	14.0









#23
 Vinyl Acetate
 Concen: 619.743 ug/l
 RT: 6.431 min Scan# 1656
 Delta R.T. -0.000 min
 Lab File: VN070665.D
 Acq: 18 Jan 2022 12:34

Tgt Ion: 43 Resp:18737374




Ion	Ratio	Lower	Upper
43	100		
86	8.7	8.5	12.7




#24
 1,1-Dichloroethane
 Concen: 108.239 ug/l
 RT: 6.385 min Scan# 1639
 Delta R.T. -0.000 min
 Lab File: VN070665.D
 Acq: 18 Jan 2022 12:34

Tgt Ion: 63 Resp: 2139489





Ion	Ratio	Lower	Upper
63	100		
98	5.9	3.3	9.9
100	3.6	2.1	6.5








#25
2-Butanone
Concen: 561.287 ug/l
RT: 7.332 min Scan# 1992
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion: 43 Resp: 5626844
Ion Ratio Lower Upper
43 100
72 23.1 21.7 32.5



#26
2,2-Dichloropropane
Concen: 107.367 ug/l
RT: 7.324 min Scan# 1989
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 77 Resp: 1662663
Ion Ratio Lower Upper
77 100
97 21.9 11.6 34.8









#27
 cis-1,2-Dichloroethene
 Concen: 108.724 ug/l
 RT: 7.324 min Scan# 1989
 Delta R.T. -0.000 min
 Lab File: VN070665.D
 Acq: 18 Jan 2022 12:34

Tgt Ion: 96 Resp: 1181956




Ion	Ratio	Lower	Upper
96	100		
61	161.1	0.0	275.0
98	64.3	0.0	129.6



#28
 Bromochloromethane
 Concen: 101.668 ug/l
 RT: 7.656 min Scan# 2113
 Delta R.T. -0.000 min
 Lab File: VN070665.D
 Acq: 18 Jan 2022 12:34

Tgt Ion: 49 Resp: 963918

Ion	Ratio	Lower	Upper
49	100		
129	1.7	0.0	4.0
130	61.7	67.8	101.8#





#29

Tetrahydrofuran

Concen: 574.009 ug/l

RT: 7.681 min Scan# 2122





Delta R.T. -0.003 min

Lab File: VN070665.D

Acq: 18 Jan 2022 12:34

Tgt Ion: 42 Resp: 3449369

Ion	Ratio	Lower	Upper
42	100		
72	37.0	36.3	54.5
71	35.4	34.7	52.1



#30

Chloroform

Concen: 105.635 ug/l

RT: 7.815 min Scan# 2172




Delta R.T. -0.003 min


Lab File: VN070665.D

Acq: 18 Jan 2022 12:34

Tgt Ion: 83 Resp: 2007351

Ion	Ratio	Lower	Upper
83	100		
85	63.9	51.9	77.9









#31
Cyclohexane
Concen: 111.097 ug/l
RT: 8.094 min Scan# 2276
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 56 Resp: 2158017




Ion	Ratio	Lower	Upper
56	100		
69	30.2	27.1	40.7
84	75.2	73.4	110.0




#32
1,1,1-Trichloroethane
Concen: 105.992 ug/l
RT: 8.013 min Scan# 2246
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 97 Resp: 1687108

Ion	Ratio	Lower	Upper
97	100		
99	64.2	50.1	75.1
61	47.0	35.2	52.8





#33

1,2-Dichloroethane-d4

Concen: 103.749 ug/l

RT: 8.434 min Scan# 2403

Delta R.T. -0.000 min

Lab File: VN070665.D





Acq: 18 Jan 2022 12:34

Tgt Ion: 65 Resp: 1258809

Ion Ratio Lower Upper

65 100

67 54.0 0.0 103.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.965 min Scan# 2601

Delta R.T. -0.000 min

Lab File: VN070665.D

Acq: 18 Jan 2022 12:34




Tgt Ion: 114 Resp: 1398268


Ion Ratio Lower Upper

114 100

63 23.6 0.0 37.8





88 16.7 0.0 27.4








#35
Dibromofluoromethane
Concen: 107.635 ug/l
RT: 8.021 min Scan# 2249
Delta R.T. 0.003 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion: 113 Resp: 915436
Ion Ratio Lower Upper
113 100
111 102.6 82.2 123.2
192 19.8 16.9 25.3



#36
1,1-Dichloropropene
Concen: 110.854 ug/l
RT: 8.220 min Scan# 2323
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 75 Resp: 1559139
Ion Ratio Lower Upper
75 100
110 32.3 18.4 55.2
77 30.3 24.4 36.6









#37
Ethyl Acetate
Concen: 111.315 ug/l
RT: 7.410 min Scan# 2021
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 43 Resp: 2045199




Ion	Ratio	Lower	Upper
43	100		
61	13.2	11.5	17.3
70	9.2	7.9	11.9




#38
Carbon Tetrachloride
Concen: 106.222 ug/l
RT: 8.206 min Scan# 2318
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 117 Resp: 1411893

Ion	Ratio	Lower	Upper
117	100		
119	95.1	76.1	114.1
121	30.5	24.2	36.2









#39
Methylcyclohexane
Concen: 116.466 ug/l
RT: 9.459 min Scan# 2785
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 83 Resp: 1994381




Ion	Ratio	Lower	Upper
83	100		
55	85.0	58.1	87.1
98	44.3	38.4	57.6




#40
Benzene
Concen: 109.174 ug/l
RT: 8.458 min Scan# 2412
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 78 Resp: 4718645

Ion	Ratio	Lower	Upper
78	100		
77	23.9	18.6	27.8









#41
Methacrylonitrile
Concen: 120.014 ug/l
RT: 7.635 min Scan# 2105
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 41 Resp: 1094701




Ion	Ratio	Lower	Upper
41	100		
39	55.1	48.1	72.1
67	67.6	68.3	102.5#
52	27.4	29.6	44.4#




#42
1,2-Dichloroethane
Concen: 101.130 ug/l
RT: 8.528 min Scan# 2438
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 62 Resp: 1624689

Ion	Ratio	Lower	Upper
62	100		
98	8.6	0.0	17.6









#43
Isopropyl Acetate
Concen: 112.621 ug/l
RT: 8.560 min Scan# 2450
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 43 Resp: 3245677




Ion	Ratio	Lower	Upper
43	100		
61	24.0	20.4	30.6
87	11.3	11.2	16.8



#44
Trichloroethene
Concen: 106.696 ug/l
RT: 9.215 min Scan# 2694
Delta R.T. 0.003 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion:130 Resp: 1090608

Ion	Ratio	Lower	Upper
130	100		
95	106.2	0.0	191.2





#45

1,2-Dichloropropane

Concen: 111.608 ug/l

RT: 9.488 min Scan# 2796

Delta R.T. -0.000 min

Lab File: VN070665.D





Acq: 18 Jan 2022 12:34

Tgt Ion: 63 Resp: 1300956

Ion Ratio Lower Upper

63 100

65 30.8 24.7 37.1



#46

Dibromomethane

Concen: 105.886 ug/l

RT: 9.574 min Scan# 2828

Delta R.T. -0.000 min

Lab File: VN070665.D

Acq: 18 Jan 2022 12:34




Tgt Ion: 93 Resp: 761047


Ion Ratio Lower Upper

93 100

95 82.4 65.9 98.9

174 95.1 84.6 127.0









#47
Bromodichloromethane
Concen: 110.190 ug/l
RT: 9.759 min Scan# 2897
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 83 Resp: 1569285




Ion	Ratio	Lower	Upper
83	100		
85	63.8	50.9	76.3
127	7.8	7.1	10.7




#48
Methyl methacrylate
Concen: 117.580 ug/l
RT: 9.558 min Scan# 2822
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 41 Resp: 1609276





Ion	Ratio	Lower	Upper
41	100		
69	74.7	72.4	108.6
39	48.9	58.4	87.6








#49
1,4-Dioxane
Concen: 2304.518 ug/l
RT: 9.569 min Scan# 2826
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion: 88 Resp: 467797
Ion Ratio Lower Upper
88 100
43 34.0 23.4 35.2
58 79.4 50.6 75.8#



#50
Toluene-d8
Concen: 109.462 ug/l
RT: 10.440 min Scan# 3151
Delta R.T. 0.003 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion: 98 Resp: 3830964
Ion Ratio Lower Upper
98 100
100 64.8 50.4 75.6








#51
4-Methyl-2-Pentanone
Concen: 573.497 ug/l
RT: 10.330 min Scan# 3110
Delta R.T. 0.003 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion: 43 Resp:10404798
Ion Ratio Lower Upper
43 100
58 37.3 30.4 45.6



#52
Toluene
Concen: 111.178 ug/l
RT: 10.505 min Scan# 3175
Delta R.T. 0.003 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion: 92 Resp: 2902253
Ion Ratio Lower Upper
92 100
91 173.5 142.0 213.0








#53
t-1,3-Dichloropropene
Concen: 118.006 ug/l
RT: 10.717 min Scan# 3254
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion: 75 Resp: 1825800
Ion Ratio Lower Upper
75 100
77 31.5 25.6 38.4



#54
cis-1,3-Dichloropropene
Concen: 118.535 ug/l
RT: 10.188 min Scan# 3057
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 75 Resp: 1992424
Ion Ratio Lower Upper
75 100
77 31.1 26.4 39.6
39 53.1 42.5 63.7









#55
1,1,2-Trichloroethane
Concen: 108.358 ug/l
RT: 10.896 min Scan# 3321
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 97 Resp: 1117732




Ion	Ratio	Lower	Upper
97	100		
83	90.0	67.8	101.6
85	57.5	45.4	68.2
99	62.2	49.4	74.0



#56
Ethyl methacrylate
Concen: 121.748 ug/l
RT: 10.759 min Scan# 3270
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 69 Resp: 2087410





Ion	Ratio	Lower	Upper
69	100		
41	73.3	51.4	77.0
39	35.5	34.6	52.0








#57
1,3-Dichloropropane
Concen: 110.524 ug/l
RT: 11.041 min Scan# 3375
Delta R.T. -0.003 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion: 76 Resp: 2022131
Ion Ratio Lower Upper
76 100
78 32.1 25.7 38.5



#58
2-Chloroethyl Vinyl ether
Concen: 717.593 ug/l
RT: 10.041 min Scan# 3002
Delta R.T. 0.003 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion: 63 Resp: 2634912
Ion Ratio Lower Upper
63 100
106 24.3 23.9 35.9








#59
2-Hexanone
Concen: 589.035 ug/l
RT: 11.084 min Scan# 3391
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion: 43 Resp: 7804442
Ion Ratio Lower Upper
43 100
58 51.7 26.9 80.7



#60
Dibromochloromethane
Concen: 113.711 ug/l
RT: 11.237 min Scan# 3448
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion: 129 Resp: 1136352
Ion Ratio Lower Upper
129 100
127 77.4 38.6 116.0








#61
1,2-Dibromoethane
Concen: 111.747 ug/l
RT: 11.344 min Scan# 3488
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion:107 Resp: 1169636
Ion Ratio Lower Upper
107 100
109 93.2 75.7 113.5



#62
4-Bromofluorobenzene
Concen: 115.397 ug/l
RT: 12.728 min Scan# 4004
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion: 95 Resp: 1452643
Ion Ratio Lower Upper
95 100
174 73.7 0.0 187.2
176 71.0 0.0 181.4








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.741 min Scan# 3636
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion:117 Resp: 1322861
Ion Ratio Lower Upper
117 100
82 60.7 42.4 63.6
119 32.0 25.4 38.2



#64
Tetrachloroethene
Concen: 99.656 ug/l
RT: 10.977 min Scan# 3351
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion:164 Resp: 957724
Ion Ratio Lower Upper
164 100
166 129.1 101.9 152.9
129 99.8 72.0 108.0
131 94.6 70.9 106.3








#65
Chlorobenzene
Concen: 108.091 ug/l
RT: 11.768 min Scan# 3646
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion:112 Resp: 3037371
Ion Ratio Lower Upper
112 100
114 32.1 25.4 38.0



#66
1,1,1,2-Tetrachloroethane
Concen: 106.014 ug/l
RT: 11.840 min Scan# 3673
Delta R.T. 0.003 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion:131 Resp: 1060539
Ion Ratio Lower Upper
131 100
133 94.9 47.9 143.7
119 65.8 32.1 96.5





#67

Ethyl Benzene

Concen: 111.114 ug/l

RT: 11.843 min Scan# 3674

Delta R.T. -0.000 min

Lab File: VN070665.D





Acq: 18 Jan 2022 12:34

Tgt Ion: 91 Resp: 5737805

Ion Ratio Lower Upper

91 100

106 29.7 25.7 38.5



#68

m/p-Xylenes

Concen: 223.309 ug/l

RT: 11.950 min Scan# 3714

Delta R.T. -0.000 min

Lab File: VN070665.D




Acq: 18 Jan 2022 12:34


Tgt Ion:106 Resp: 4200500

Ion Ratio Lower Upper

106 100





91 213.2 160.6 240.8








#69
o-Xylene
Concen: 112.530 ug/l
RT: 12.277 min Scan# 3836
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion:106 Resp: 2102107
Ion Ratio Lower Upper
106 100
91 225.7 105.8 317.3



#70
Styrene
Concen: 119.059 ug/l
RT: 12.291 min Scan# 3841
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion:104 Resp: 3526920
Ion Ratio Lower Upper
104 100
78 54.0 38.7 58.1
103 55.7 45.0 67.6








#71
Bromoform
Concen: 117.071 ug/l
RT: 12.454 min Scan# 3902
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion:173 Resp: 867833
Ion Ratio Lower Upper
173 100
175 49.0 24.2 72.6
254 0.0 0.0 0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.672 min Scan# 4356
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion:152 Resp: 567560
Ion Ratio Lower Upper
152 100
115 60.7 27.5 82.5
150 193.4 0.0 344.4









#73
Isopropylbenzene
Concen: 105.055 ug/l
RT: 12.575 min Scan# 3947
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion:105 Resp: 5523223




Ion	Ratio	Lower	Upper
105	100		
120	25.7	13.4	40.1




#74
N-ethyl acetate
Concen: 111.122 ug/l
RT: 12.385 min Scan# 3876
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 43 Resp: 2470801





Ion	Ratio	Lower	Upper
43	100		
70	37.1	33.4	50.0
55	22.8	19.3	28.9
61	22.3	18.5	27.7








#75
1,1,2,2-Tetrachloroethane
Concen: 97.541 ug/l
RT: 12.825 min Scan# 4040
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion: 83 Resp: 1730408
Ion Ratio Lower Upper
83 100
131 9.7 5.5 16.5
85 64.0 32.4 97.2



#76
1,2,3-Trichloropropane
Concen: 132.790 ug/l
RT: 12.878 min Scan# 4060
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion: 75 Resp: 2231718
Ion Ratio Lower Upper
75 100
77 0.0 82.6 247.8#








#77
Bromobenzene
Concen: 103.688 ug/l
RT: 12.859 min Scan# 4053
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion:156 Resp: 1248978
Ion Ratio Lower Upper
156 100
77 235.0 90.6 271.8
158 96.5 48.9 146.7



#78
n-propylbenzene
Concen: 111.369 ug/l
RT: 12.918 min Scan# 4075
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion: 91 Resp: 6517578
Ion Ratio Lower Upper
91 100
120 21.8 11.9 35.9








#79
2-Chlorotoluene
Concen: 104.575 ug/l
RT: 13.004 min Scan# 4107
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 91 Resp: 3917481
Ion Ratio Lower Upper
91 100
126 31.9 18.3 54.8



#80
1,3,5-Trimethylbenzene
Concen: 108.668 ug/l
RT: 13.058 min Scan# 4127
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion: 105 Resp: 4586276
Ion Ratio Lower Upper
105 100
120 47.2 24.6 73.6








#81
trans-1,4-Dichloro-2-butene
Concen: 124.540 ug/l
RT: 12.623 min Scan# 3965
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion: 75 Resp: 581194
Ion Ratio Lower Upper
75 100
53 100.1 90.5 135.7
89 44.5 38.2 57.4



#82
4-Chlorotoluene
Concen: 108.955 ug/l
RT: 13.101 min Scan# 4143
Delta R.T. -0.003 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion: 91 Resp: 3835336
Ion Ratio Lower Upper
91 100
126 31.3 18.2 54.6








#83
tert-Butylbenzene
Concen: 107.867 ug/l
RT: 13.321 min Scan# 4225
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion:119 Resp: 4011118
Ion Ratio Lower Upper
119 100
91 68.9 31.5 94.5
134 25.3 12.4 37.4



#84
1,2,4-Trimethylbenzene
Concen: 110.185 ug/l
RT: 13.366 min Scan# 4242
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion:105 Resp: 4480247
Ion Ratio Lower Upper
105 100
120 43.8 22.8 68.4








#85
 sec-Butylbenzene
 Concen: 112.240 ug/l
 RT: 13.500 min Scan# 4292
 Delta R.T. 0.003 min
 Lab File: VN070665.D
 Acq: 18 Jan 2022 12:34


Tgt Ion:105 Resp: 5671976
 Ion Ratio Lower Upper
 105 100
 134 19.3 10.4 31.1



#86
 p-Isopropyltoluene
 Concen: 114.277 ug/l
 RT: 13.613 min Scan# 4334
 Delta R.T. -0.000 min
 Lab File: VN070665.D
 Acq: 18 Jan 2022 12:34





Tgt Ion:119 Resp: 4539494
 Ion Ratio Lower Upper
 119 100
 134 25.6 13.6 40.7
 91 24.8 11.4 34.1








#87
1,3-Dichlorobenzene
Concen: 104.249 ug/l
RT: 13.613 min Scan# 4334
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion:146 Resp: 2112428
Ion Ratio Lower Upper
146 100
111 42.3 18.6 55.6
148 64.0 31.6 95.0



#88
1,4-Dichlorobenzene
Concen: 103.375 ug/l
RT: 13.694 min Scan# 4364
Delta R.T. 0.003 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion:146 Resp: 2068111
Ion Ratio Lower Upper
146 100
111 40.7 18.3 54.8
148 63.6 31.8 95.4








#89
n-Butylbenzene
Concen: 118.223 ug/l
RT: 13.940 min Scan# 4456
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion: 91 Resp: 3921354
Ion Ratio Lower Upper
91 100
92 52.1 25.4 76.0
134 24.4 13.6 40.8



#90
Hexachloroethane
Concen: 115.702 ug/l
RT: 14.209 min Scan# 4556
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion: 117 Resp: 794802
Ion Ratio Lower Upper
117 100
201 80.3 43.9 131.6








#91
1,2-Dichlorobenzene
Concen: 101.184 ug/l
RT: 13.986 min Scan# 4473
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion:146 Resp: 2009037
Ion Ratio Lower Upper
146 100
111 43.9 19.1 57.3
148 63.8 32.1 96.3



#92
1,2-Dibromo-3-Chloropropane
Concen: 108.822 ug/l
RT: 14.600 min Scan# 4702
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34





Tgt Ion: 75 Resp: 332206
Ion Ratio Lower Upper
75 100
155 78.4 46.9 140.7
157 100.0 60.6 181.7








#93
1,2,4-Trichlorobenzene
Concen: 115.003 ug/l
RT: 15.263 min Scan# 4949
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34


Tgt Ion:180 Resp: 1083602
Ion Ratio Lower Upper
180 100
182 95.4 48.3 144.9
145 30.8 14.5 43.5



#94
Hexachlorobutadiene
Concen: 106.239 ug/l
RT: 15.367 min Scan# 4988
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion:225 Resp: 586601
Ion Ratio Lower Upper
225 100
223 63.1 31.9 95.7
227 63.8 32.1 96.3









#95
Naphthalene
Concen: 119.941 ug/l
RT: 15.504 min Scan# 5039
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion:128 Resp: 3063778

Ion	Ratio	Lower	Upper
128	100		
127	13.1	10.5	15.7
129	10.8	8.7	13.1



#96
1,2,3-Trichlorobenzene
Concen: 112.661 ug/l
RT: 15.697 min Scan# 5111
Delta R.T. -0.000 min
Lab File: VN070665.D
Acq: 18 Jan 2022 12:34

Tgt Ion:180 Resp: 1036364

Ion	Ratio	Lower	Upper
180	100		
182	95.0	47.4	142.2
145	32.7	15.0	45.1

