

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011822\
 Data File : VN070670.D
 Acq On : 18 Jan 2022 17:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Quant Time: Jan 19 01:10:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.083	168	891942	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1423869	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1333030	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	581826	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	557803	43.637	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	87.280%
35) Dibromofluoromethane	8.019	113	406100	45.979	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.960%
50) Toluene-d8	10.440	98	1699591	45.678	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	91.360%
62) 4-Bromofluorobenzene	12.728	95	630108	47.523	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.040%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.073	85	468097	45.416	ug/l	98
3) Chloromethane	2.303	50	660373	43.833	ug/l	99
4) Vinyl Chloride	2.448	62	638743	44.577	ug/l	99
5) Bromomethane	2.848	94	362406	48.560	ug/l	98
6) Chloroethane	3.017	64	398360	44.549	ug/l	96
7) Trichlorofluoromethane	3.376	101	771606	47.138	ug/l	99
8) Diethyl Ether	3.824	74	319293	44.653	ug/l	80
9) 1,1,2-Trichlorotrifluo...	4.207	101	440146	44.489	ug/l	95
10) Methyl Iodide	4.419	142	590431	46.281	ug/l	100
11) Tert butyl alcohol	5.374	59	515157	223.378	ug/l	98
12) 1,1-Dichloroethene	4.181	96	436277	44.214	ug/l	90
13) Acrolein	4.039	56	695351	210.232	ug/l	100
14) Allyl chloride	4.838	41	1011738	47.059	ug/l #	84
15) Acrylonitrile	5.551	53	1675437	228.369	ug/l	99
16) Acetone	4.285	43	1799010	202.767	ug/l	98
17) Carbon Disulfide	4.527	76	1236312	46.133	ug/l	98
18) Methyl Acetate	4.854	43	758140	43.450	ug/l	92
19) Methyl tert-butyl Ether	5.616	73	1658944	47.062	ug/l	93
20) Methylene Chloride	5.095	84	524666	44.894	ug/l	86
21) trans-1,2-Dichloroethene	5.597	96	467704	47.095	ug/l	85
22) Diisopropyl ether	6.495	45	2006816	47.093	ug/l #	91
23) Vinyl Acetate	6.431	43	8650223	239.464	ug/l	95
24) 1,1-Dichloroethane	6.385	63	987285	46.316	ug/l	98
25) 2-Butanone	7.332	43	2478259	221.979	ug/l	92
26) 2,2-Dichloropropane	7.324	77	770075	47.119	ug/l	97
27) cis-1,2-Dichloroethene	7.321	96	546053	46.286	ug/l	86
28) Bromochloromethane	7.657	49	437603	45.724	ug/l #	75
29) Tetrahydrofuran	7.683	42	1575592	230.484	ug/l	87
30) Chloroform	7.815	83	954631	45.819	ug/l	100
31) Cyclohexane	8.094	56	1002168	44.522	ug/l	87
32) 1,1,1-Trichloroethane	8.013	97	776473	46.930	ug/l	98
36) 1,1-Dichloropropene	8.220	75	722539	48.594	ug/l	96
37) Ethyl Acetate	7.410	43	934001	47.315	ug/l	97
38) Carbon Tetrachloride	8.204	117	639523	48.347	ug/l	99
39) Methylcyclohexane	9.459	83	912061	49.687	ug/l	89
40) Benzene	8.458	78	2201222	47.778	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011822\
 Data File : VN070670.D
 Acq On : 18 Jan 2022 17:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Quant Time: Jan 19 01:10:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.635	41	467266	46.527	ug/l	89
42) 1,2-Dichloroethane	8.528	62	766851	48.003	ug/l	99
43) Isopropyl Acetate	8.558	43	1449188	48.064	ug/l #	96
44) Trichloroethene	9.215	130	502997	48.158	ug/l	90
45) 1,2-Dichloropropane	9.488	63	602521	48.066	ug/l	100
46) Dibromomethane	9.574	93	351964	48.289	ug/l	94
47) Bromodichloromethane	9.759	83	719109	48.265	ug/l	100
48) Methyl methacrylate	9.558	41	721669	50.241	ug/l #	77
49) 1,4-Dioxane	9.566	88	220401	1021.499	ug/l #	83
51) 4-Methyl-2-Pentanone	10.328	43	4660483	242.236	ug/l	98
52) Toluene	10.502	92	1349566	48.970	ug/l	97
53) t-1,3-Dichloropropene	10.717	75	818997	45.554	ug/l	98
54) cis-1,3-Dichloropropene	10.188	75	897111	50.208	ug/l	98
55) 1,1,2-Trichloroethane	10.896	97	515350	47.601	ug/l	97
56) Ethyl methacrylate	10.760	69	908275	44.382	ug/l	89
57) 1,3-Dichloropropane	11.044	76	930656	48.558	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.038	63	1124590	218.245	ug/l	90
59) 2-Hexanone	11.084	43	3447827	246.686	ug/l	97
60) Dibromochloromethane	11.237	129	509074	50.410	ug/l	100
61) 1,2-Dibromoethane	11.344	107	524041	48.210	ug/l	99
64) Tetrachloroethene	10.974	164	460598	47.126	ug/l	94
65) Chlorobenzene	11.768	112	1381395	47.830	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.838	131	479628	48.225	ug/l	99
67) Ethyl Benzene	11.843	91	2631017	49.135	ug/l	96
68) m/p-Xylenes	11.953	106	1939446	99.562	ug/l	92
69) o-Xylene	12.278	106	956186	49.463	ug/l	92
70) Styrene	12.291	104	1589326	51.391	ug/l	96
71) Bromoform	12.455	173	372645	49.603	ug/l #	100
73) Isopropylbenzene	12.578	105	2516329	45.715	ug/l	98
74) N-amyl acetate	12.385	43	1072529	47.656	ug/l	96
75) 1,1,2,2-Tetrachloroethane	12.825	83	815472	43.868	ug/l	98
76) 1,2,3-Trichloropropane	12.878	75	1014808	65.762	ug/l #	1
77) Bromobenzene	12.860	156	566610	45.912	ug/l	75
78) n-propylbenzene	12.919	91	2974833	48.326	ug/l	96
79) 2-Chlorotoluene	13.004	91	1794896	46.094	ug/l	92
80) 1,3,5-Trimethylbenzene	13.058	105	2097955	47.132	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.624	75	244666	47.234	ug/l	94
82) 4-Chlorotoluene	13.104	91	1755657	48.255	ug/l	91
83) tert-Butylbenzene	13.321	119	1804251	46.564	ug/l	93
84) 1,2,4-Trimethylbenzene	13.366	105	2054390	48.015	ug/l	98
85) sec-Butylbenzene	13.501	105	2570251	48.793	ug/l	97
86) p-Isopropyltoluene	13.613	119	2070232	50.029	ug/l	96
87) 1,3-Dichlorobenzene	13.613	146	993178	48.378	ug/l	97
88) 1,4-Dichlorobenzene	13.694	146	965863	47.275	ug/l	97
89) n-Butylbenzene	13.940	91	1746301	51.154	ug/l	96
90) Hexachloroethane	14.209	117	345797	47.805	ug/l	91
91) 1,2-Dichlorobenzene	13.986	146	946899	46.684	ug/l	96
92) 1,2-Dibromo-3-Chloropr...	14.600	75	139078	44.105	ug/l	82
93) 1,2,4-Trichlorobenzene	15.263	180	471340	44.270	ug/l	98
94) Hexachlorobutadiene	15.367	225	298768	49.735	ug/l	100
95) Naphthalene	15.504	128	1177443	40.113	ug/l	100
96) 1,2,3-Trichlorobenzene	15.700	180	447081	43.928	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011822\
Data File : VN070670.D
Acq On : 18 Jan 2022 17:02
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1


Quant Time: Jan 19 01:10:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed





Quant Time: Jan 19 01:10:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.083 min Scan# 2272
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion:168 Resp: 891942
Ion Ratio Lower Upper
168 100
99 59.0 41.3 61.9



#2
Dichlorodifluoromethane
Concen: 45.416 ug/l
RT: 2.073 min Scan# 31
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion: 85 Resp: 468097
Ion Ratio Lower Upper
85 100
87 31.7 16.3 48.9








#3
Chloromethane
Concen: 43.833 ug/l
RT: 2.303 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 50 Resp: 660373
Ion Ratio Lower Upper
50 100
52 32.2 25.4 38.0



#4
Vinyl Chloride
Concen: 44.577 ug/l
RT: 2.448 min Scan# 171
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion: 62 Resp: 638743
Ion Ratio Lower Upper
62 100
64 31.5 25.8 38.6








#5
Bromomethane
Concen: 48.560 ug/l
RT: 2.848 min Scan# 320
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 94 Resp: 362406
Ion Ratio Lower Upper
94 100
96 93.8 73.8 110.8



#6
Chloroethane
Concen: 44.549 ug/l
RT: 3.017 min Scan# 383
Delta R.T. -0.002 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion: 64 Resp: 398360
Ion Ratio Lower Upper
64 100
66 30.9 26.3 39.5








#7
Trichlorofluoromethane
Concen: 47.138 ug/l
RT: 3.376 min Scan# 517
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion:101 Resp: 771606
Ion Ratio Lower Upper
101 100
103 63.4 50.1 75.1



#8
Diethyl Ether
Concen: 44.653 ug/l
RT: 3.824 min Scan# 684
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion: 74 Resp: 319293
Ion Ratio Lower Upper
74 100
45 113.6 47.0 141.2








#9
1,1,2-Trichlorotrifluoroethane
Concen: 44.489 ug/l
RT: 4.207 min Scan# 827
Delta R.T. -0.003 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion:	101	Resp:	440146
Ion	Ratio	Lower	Upper
101	100		
85	44.8	37.2	55.8
151	76.2	65.9	98.9



#10
Methyl Iodide
Concen: 46.281 ug/l
RT: 4.419 min Scan# 906
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion:	142	Resp:	590431
Ion	Ratio	Lower	Upper
142	100		
127	43.8	35.4	53.0
141	14.5	11.6	17.4








#11
Tert butyl alcohol
Concen: 223.378 ug/l
RT: 5.374 min Scan# 1262
Delta R.T. -0.003 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 59 Resp: 515157
Ion Ratio Lower Upper
59 100
57 10.0 8.7 13.1



#12
1,1-Dichloroethene
Concen: 44.214 ug/l
RT: 4.181 min Scan# 817
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion: 96 Resp: 436277
Ion Ratio Lower Upper
96 100
61 180.6 131.0 196.6
98 61.8 52.4 78.6








#13
Acrolein
Concen: 210.232 ug/l
RT: 4.039 min Scan# 764
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 56 Resp: 695351
Ion Ratio Lower Upper
56 100
55 70.4 56.6 85.0



#14
Allyl chloride
Concen: 47.059 ug/l
RT: 4.838 min Scan# 1062
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 41 Resp: 1011738
Ion Ratio Lower Upper
41 100
39 61.0 61.2 91.8#
76 26.7 26.0 39.0









#15
Acrylonitrile
Concen: 228.369 ug/l
RT: 5.551 min Scan# 1328
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 53 Resp: 1675437




Ion	Ratio	Lower	Upper
53	100		
52	83.0	66.2	99.2
51	35.7	29.6	44.4




#16
Acetone
Concen: 202.767 ug/l
RT: 4.285 min Scan# 856
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 43 Resp: 1799010





Ion	Ratio	Lower	Upper
43	100		
58	30.2	23.1	34.7








#17
Carbon Disulfide
Concen: 46.133 ug/l
RT: 4.527 min Scan# 946
Delta R.T. -0.002 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 76 Resp: 1236312
Ion Ratio Lower Upper
76 100
78 8.7 7.5 11.3



#18
Methyl Acetate
Concen: 43.450 ug/l
RT: 4.854 min Scan# 1068
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 43 Resp: 758140
Ion Ratio Lower Upper
43 100
74 20.3 19.4 29.2









#19

Methyl tert-butyl Ether
Concen: 47.062 ug/l
RT: 5.616 min Scan# 1352
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02




Tgt Ion: 73 Resp: 1658944
Ion Ratio Lower Upper
73 100
57 24.2 16.8 25.2




#20

Methylene Chloride
Concen: 44.894 ug/l
RT: 5.095 min Scan# 1158
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion: 84 Resp: 524666
Ion Ratio Lower Upper
84 100
49 139.6 93.5 140.3
51 42.6 29.4 44.2
86 63.4 53.0 79.4








#21
trans-1,2-Dichloroethene
Concen: 47.095 ug/l
RT: 5.597 min Scan# 1345
Delta R.T. 0.003 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 96 Resp: 467704
Ion Ratio Lower Upper
96 100
61 158.5 105.8 158.6
98 64.4 52.0 78.0



#22
Diisopropyl ether
Concen: 47.093 ug/l
RT: 6.495 min Scan# 1680
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 45 Resp: 2006816
Ion Ratio Lower Upper
45 100
43 56.7 51.7 77.5
87 23.1 23.2 34.8#
59 10.6 9.4 14.0





#23

Vinyl Acetate

Concen: 239.464 ug/l

RT: 6.431 min Scan# 1656





Delta R.T. -0.000 min

Lab File: VN070670.D

Acq: 18 Jan 2022 17:02

Tgt Ion: 43 Resp: 8650223

Ion	Ratio	Lower	Upper
43	100		
86	8.6	8.5	12.7



#24

1,1-Dichloroethane

Concen: 46.316 ug/l

RT: 6.385 min Scan# 1639




Delta R.T. 0.000 min


Lab File: VN070670.D

Acq: 18 Jan 2022 17:02

Tgt Ion: 63 Resp: 987285





Ion	Ratio	Lower	Upper
63	100		
98	6.1	3.3	9.9
100	3.8	2.1	6.5








#25
2-Butanone
Concen: 221.979 ug/l
RT: 7.332 min Scan# 1992
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 43 Resp: 2478259
Ion Ratio Lower Upper
43 100
72 22.9 21.7 32.5



#26
2,2-Dichloropropane
Concen: 47.119 ug/l
RT: 7.324 min Scan# 1989
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion: 77 Resp: 770075
Ion Ratio Lower Upper
77 100
97 21.9 11.6 34.8








#27
cis-1,2-Dichloroethene
Concen: 46.286 ug/l
RT: 7.321 min Scan# 1988
Delta R.T. -0.003 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 96 Resp: 546053
Ion Ratio Lower Upper
96 100
61 161.9 0.0 275.0
98 65.0 0.0 129.6



#28
Bromochloromethane
Concen: 45.724 ug/l
RT: 7.657 min Scan# 2113
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 49 Resp: 437603
Ion Ratio Lower Upper
49 100
129 1.6 0.0 4.0
130 61.1 67.8 101.8#









#29
Tetrahydrofuran
Concen: 230.484 ug/l
RT: 7.683 min Scan# 2123
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 42 Resp: 1575592




Ion	Ratio	Lower	Upper
42	100		
72	36.6	36.3	54.5
71	35.2	34.7	52.1




#30
Chloroform
Concen: 45.819 ug/l
RT: 7.815 min Scan# 2172
Delta R.T. -0.003 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 83 Resp: 954631

Ion	Ratio	Lower	Upper
83	100		
85	65.0	51.9	77.9









#31
Cyclohexane
Concen: 44.522 ug/l
RT: 8.094 min Scan# 2276
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 56 Resp: 1002168




Ion	Ratio	Lower	Upper
56	100		
69	31.4	27.1	40.7
84	76.8	73.4	110.0




#32
1,1,1-Trichloroethane
Concen: 46.930 ug/l
RT: 8.013 min Scan# 2246
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 97 Resp: 776473





Ion	Ratio	Lower	Upper
97	100		
99	63.2	50.1	75.1
61	47.2	35.2	52.8








#33
1,2-Dichloroethane-d4
Concen: 43.637 ug/l
RT: 8.434 min Scan# 2403
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 65 Resp: 557803
Ion Ratio Lower Upper
65 100
67 53.5 0.0 103.0



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.965 min Scan# 2601
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion: 114 Resp: 1423869
Ion Ratio Lower Upper
114 100
63 23.5 0.0 37.8
88 17.0 0.0 27.4








#35
Dibromofluoromethane
Concen: 45.979 ug/l
RT: 8.019 min Scan# 2248
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion:113 Resp: 406100
Ion Ratio Lower Upper
113 100
111 102.8 82.2 123.2
192 19.8 16.9 25.3



#36
1,1-Dichloropropene
Concen: 48.594 ug/l
RT: 8.220 min Scan# 2323
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 75 Resp: 722539
Ion Ratio Lower Upper
75 100
110 33.0 18.4 55.2
77 30.7 24.4 36.6









#37
Ethyl Acetate
Concen: 47.315 ug/l
RT: 7.410 min Scan# 2021
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 43 Resp: 934001




Ion	Ratio	Lower	Upper
43	100		
61	13.1	11.5	17.3
70	9.0	7.9	11.9




#38
Carbon Tetrachloride
Concen: 48.347 ug/l
RT: 8.204 min Scan# 2317
Delta R.T. -0.002 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 117 Resp: 639523

Ion	Ratio	Lower	Upper
117	100		
119	96.4	76.1	114.1
121	30.8	24.2	36.2





#39

Methylcyclohexane

Concen: 49.687 ug/l

RT: 9.459 min Scan# 2785

Delta R.T. -0.000 min

Lab File: VN070670.D

Acq: 18 Jan 2022 17:02





Tgt Ion: 83 Resp: 912061

Ion Ratio Lower Upper

83 100

55 85.5 58.1 87.1

98 45.6 38.4 57.6



#40

Benzene

Concen: 47.778 ug/l

RT: 8.458 min Scan# 2412

Delta R.T. -0.001 min

Lab File: VN070670.D




Acq: 18 Jan 2022 17:02


Tgt Ion: 78 Resp: 2201222

Ion Ratio Lower Upper

78 100





77 23.3 18.6 27.8








#41
Methacrylonitrile
Concen: 46.527 ug/l
RT: 7.635 min Scan# 2105
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion:	41	Resp:	467266
Ion	Ratio	Lower	Upper
41	100		
39	57.7	48.1	72.1
67	71.3	68.3	102.5
52	30.1	29.6	44.4



#42
1,2-Dichloroethane
Concen: 48.003 ug/l
RT: 8.528 min Scan# 2438
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion:	62	Resp:	766851
Ion	Ratio	Lower	Upper
62	100		
98	8.4	0.0	17.6









#43
Isopropyl Acetate
Concen: 48.064 ug/l
RT: 8.558 min Scan# 2449
Delta R.T. -0.002 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 43 Resp: 1449188




Ion	Ratio	Lower	Upper
43	100		
61	24.1	20.4	30.6
87	11.0	11.2	16.8#




#44
Trichloroethene
Concen: 48.158 ug/l
RT: 9.215 min Scan# 2694
Delta R.T. 0.003 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion:130 Resp: 502997





Ion	Ratio	Lower	Upper
130	100		
95	105.4	0.0	191.2








#45
1,2-Dichloropropane
Concen: 48.066 ug/l
RT: 9.488 min Scan# 2796
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 63 Resp: 602521
Ion Ratio Lower Upper
63 100
65 31.0 24.7 37.1



#46
Dibromomethane
Concen: 48.289 ug/l
RT: 9.574 min Scan# 2828
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 93 Resp: 351964
Ion Ratio Lower Upper
93 100
95 83.2 65.9 98.9
174 95.3 84.6 127.0









#47
Bromodichloromethane
Concen: 48.265 ug/l
RT: 9.759 min Scan# 2897
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 83 Resp: 719109




Ion	Ratio	Lower	Upper
83	100		
85	63.8	50.9	76.3
127	8.1	7.1	10.7




#48
Methyl methacrylate
Concen: 50.241 ug/l
RT: 9.558 min Scan# 2822
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 41 Resp: 721669





Ion	Ratio	Lower	Upper
41	100		
69	73.9	72.4	108.6
39	48.4	58.4	87.6








#49
1,4-Dioxane
Concen: 1021.499 ug/l
RT: 9.566 min Scan# 2825
Delta R.T. -0.003 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 88 Resp: 220401
Ion Ratio Lower Upper
88 100
43 33.7 23.4 35.2
58 80.0 50.6 75.8#



#50
Toluene-d8
Concen: 45.678 ug/l
RT: 10.440 min Scan# 3151
Delta R.T. 0.002 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion: 98 Resp: 1699591
Ion Ratio Lower Upper
98 100
100 65.2 50.4 75.6








#51
4-Methyl-2-Pentanone
Concen: 242.236 ug/l
RT: 10.328 min Scan# 3109
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 43 Resp: 4660483
Ion Ratio Lower Upper
43 100
58 36.8 30.4 45.6



#52
Toluene
Concen: 48.970 ug/l
RT: 10.502 min Scan# 3174
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion: 92 Resp: 1349566
Ion Ratio Lower Upper
92 100
91 172.9 142.0 213.0








#53
t-1,3-Dichloropropene
Concen: 45.554 ug/l
RT: 10.717 min Scan# 3254
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 75 Resp: 818997
Ion Ratio Lower Upper
75 100
77 31.0 25.6 38.4



#54
cis-1,3-Dichloropropene
Concen: 50.208 ug/l
RT: 10.188 min Scan# 3057
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 75 Resp: 897111
Ion Ratio Lower Upper
75 100
77 31.1 26.4 39.6
39 52.8 42.5 63.7









#55
1,1,2-Trichloroethane
Concen: 47.601 ug/l
RT: 10.896 min Scan# 3321
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 97 Resp: 515350




Ion	Ratio	Lower	Upper
97	100		
83	89.7	67.8	101.6
85	57.6	45.4	68.2
99	61.5	49.4	74.0




#56
Ethyl methacrylate
Concen: 44.382 ug/l
RT: 10.760 min Scan# 3270
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 69 Resp: 908275





Ion	Ratio	Lower	Upper
69	100		
41	73.6	51.4	77.0
39	37.5	34.6	52.0








#57
1,3-Dichloropropane
Concen: 48.558 ug/l
RT: 11.044 min Scan# 3376
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 76 Resp: 930656
Ion Ratio Lower Upper
76 100
78 32.2 25.7 38.5



#58
2-Chloroethyl Vinyl ether
Concen: 218.245 ug/l
RT: 10.038 min Scan# 3001
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion: 63 Resp: 1124590
Ion Ratio Lower Upper
63 100
106 24.3 23.9 35.9








#59
2-Hexanone
Concen: 246.686 ug/l
RT: 11.084 min Scan# 3391
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 43 Resp: 3447827
Ion Ratio Lower Upper
43 100
58 51.5 26.9 80.7



#60
Dibromochloromethane
Concen: 50.410 ug/l
RT: 11.237 min Scan# 3448
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion:129 Resp: 509074
Ion Ratio Lower Upper
129 100
127 77.0 38.6 116.0








#61
1,2-Dibromoethane
Concen: 48.210 ug/l
RT: 11.344 min Scan# 3488
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion:107 Resp: 524041
Ion Ratio Lower Upper
107 100
109 93.4 75.7 113.5



#62
4-Bromofluorobenzene
Concen: 47.523 ug/l
RT: 12.728 min Scan# 4004
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion: 95 Resp: 630108
Ion Ratio Lower Upper
95 100
174 74.5 0.0 187.2
176 72.1 0.0 181.4








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.744 min Scan# 3637
Delta R.T. 0.003 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion:	117	Resp:	1333030
Ion	Ratio	Lower	Upper
117	100		
82	60.1	42.4	63.6
119	31.9	25.4	38.2



#64
Tetrachloroethene
Concen: 47.126 ug/l
RT: 10.974 min Scan# 3350
Delta R.T. -0.003 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion:	164	Resp:	460598
Ion	Ratio	Lower	Upper
164	100		
166	128.9	101.9	152.9
129	100.2	72.0	108.0
131	94.6	70.9	106.3








#65
Chlorobenzene
Concen: 47.830 ug/l
RT: 11.768 min Scan# 3646
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion:112 Resp: 1381395
Ion Ratio Lower Upper
112 100
114 32.0 25.4 38.0



#66
1,1,1,2-Tetrachloroethane
Concen: 48.225 ug/l
RT: 11.838 min Scan# 3672
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion:131 Resp: 479628
Ion Ratio Lower Upper
131 100
133 95.9 47.9 143.7
119 65.8 32.1 96.5








#67
Ethyl Benzene
Concen: 49.135 ug/l
RT: 11.843 min Scan# 3674
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 91 Resp: 2631017
Ion Ratio Lower Upper
91 100
106 29.8 25.7 38.5



#68
m/p-Xylenes
Concen: 99.562 ug/l
RT: 11.953 min Scan# 3715
Delta R.T. 0.003 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion:106 Resp: 1939446
Ion Ratio Lower Upper
106 100
91 213.4 160.6 240.8








#69
o-Xylene
Concen: 49.463 ug/l
RT: 12.278 min Scan# 3836
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion:106 Resp: 956186
Ion Ratio Lower Upper
106 100
91 224.4 105.8 317.3



#70
Styrene
Concen: 51.391 ug/l
RT: 12.291 min Scan# 3841
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion:104 Resp: 1589326
Ion Ratio Lower Upper
104 100
78 53.6 38.7 58.1
103 55.6 45.0 67.6









#71
Bromoform
Concen: 49.603 ug/l
RT: 12.455 min Scan# 3902
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion:173 Resp: 372645




Ion	Ratio	Lower	Upper
173	100		
175	48.2	24.2	72.6
254	0.0	0.0	0.0




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.672 min Scan# 4356
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion:152 Resp: 581826





Ion	Ratio	Lower	Upper
152	100		
115	62.0	27.5	82.5
150	174.9	0.0	344.4








#73
Isopropylbenzene
Concen: 45.715 ug/l
RT: 12.578 min Scan# 3948
Delta R.T. 0.003 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion:105 Resp: 2516329
Ion Ratio Lower Upper
105 100
120 25.5 13.4 40.1



#74
N-ethyl acetate
Concen: 47.656 ug/l
RT: 12.385 min Scan# 3876
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion: 43 Resp: 1072529
Ion Ratio Lower Upper
43 100
70 37.6 33.4 50.0
55 23.7 19.3 28.9
61 22.4 18.5 27.7








#75
1,1,2,2-Tetrachloroethane
Concen: 43.868 ug/l
RT: 12.825 min Scan# 4040
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 83 Resp: 815472
Ion Ratio Lower Upper
83 100
131 9.6 5.5 16.5
85 63.7 32.4 97.2



#76
1,2,3-Trichloropropane
Concen: 65.762 ug/l
RT: 12.878 min Scan# 4060
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion: 75 Resp: 1014808
Ion Ratio Lower Upper
75 100
77 0.0 82.6 247.8#








#77
Bromobenzene
Concen: 45.912 ug/l
RT: 12.860 min Scan# 4053
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion:156 Resp: 566610
Ion Ratio Lower Upper
156 100
77 234.1 90.6 271.8
158 96.7 48.9 146.7



#78
n-propylbenzene
Concen: 48.326 ug/l
RT: 12.919 min Scan# 4075
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion: 91 Resp: 2974833
Ion Ratio Lower Upper
91 100
120 21.7 11.9 35.9








#79
2-Chlorotoluene
Concen: 46.094 ug/l
RT: 13.004 min Scan# 4107
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 91 Resp: 1794896
Ion Ratio Lower Upper
91 100
126 31.5 18.3 54.8



#80
1,3,5-Trimethylbenzene
Concen: 47.132 ug/l
RT: 13.058 min Scan# 4127
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion:105 Resp: 2097955
Ion Ratio Lower Upper
105 100
120 47.3 24.6 73.6








#81
trans-1,4-Dichloro-2-butene
Concen: 47.234 ug/l
RT: 12.624 min Scan# 3965
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion: 75 Resp: 244666
Ion Ratio Lower Upper
75 100
53 106.4 90.5 135.7
89 43.8 38.2 57.4



#82
4-Chlorotoluene
Concen: 48.255 ug/l
RT: 13.104 min Scan# 4144
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion: 91 Resp: 1755657
Ion Ratio Lower Upper
91 100
126 30.9 18.2 54.6








#83
tert-Butylbenzene
Concen: 46.564 ug/l
RT: 13.321 min Scan# 4225
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion:119 Resp: 1804251
Ion Ratio Lower Upper
119 100
91 69.3 31.5 94.5
134 23.7 12.4 37.4



#84
1,2,4-Trimethylbenzene
Concen: 48.015 ug/l
RT: 13.366 min Scan# 4242
Delta R.T. -0.001 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion:105 Resp: 2054390
Ion Ratio Lower Upper
105 100
120 44.2 22.8 68.4








#85
 sec-Butylbenzene
 Concen: 48.793 ug/l
 RT: 13.501 min Scan# 4292
 Delta R.T. 0.003 min
 Lab File: VN070670.D
 Acq: 18 Jan 2022 17:02


Tgt Ion:105 Resp: 2570251
 Ion Ratio Lower Upper
 105 100
 134 19.4 10.4 31.1



#86
 p-Isopropyltoluene
 Concen: 50.029 ug/l
 RT: 13.613 min Scan# 4334
 Delta R.T. 0.000 min
 Lab File: VN070670.D
 Acq: 18 Jan 2022 17:02





Tgt Ion:119 Resp: 2070232
 Ion Ratio Lower Upper
 119 100
 134 25.4 13.6 40.7
 91 24.9 11.4 34.1








#87
1,3-Dichlorobenzene
Concen: 48.378 ug/l
RT: 13.613 min Scan# 4334
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion:146 Resp: 993178
Ion Ratio Lower Upper
146 100
111 42.3 18.6 55.6
148 63.5 31.6 95.0



#88
1,4-Dichlorobenzene
Concen: 47.275 ug/l
RT: 13.694 min Scan# 4364
Delta R.T. 0.003 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion:146 Resp: 965863
Ion Ratio Lower Upper
146 100
111 40.9 18.3 54.8
148 63.5 31.8 95.4









#89
n-Butylbenzene
Concen: 51.154 ug/l
RT: 13.940 min Scan# 4456
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 91 Resp: 1746301




Ion	Ratio	Lower	Upper
91	100		
92	52.3	25.4	76.0
134	24.0	13.6	40.8




#90
Hexachloroethane
Concen: 47.805 ug/l
RT: 14.209 min Scan# 4556
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion: 117 Resp: 345797





Ion	Ratio	Lower	Upper
117	100		
201	79.7	43.9	131.6








#91
1,2-Dichlorobenzene
Concen: 46.684 ug/l
RT: 13.986 min Scan# 4473
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion:146 Resp: 946899
Ion Ratio Lower Upper
146 100
111 43.8 19.1 57.3
148 63.9 32.1 96.3



#92
1,2-Dibromo-3-Chloropropane
Concen: 44.105 ug/l
RT: 14.600 min Scan# 4702
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02





Tgt Ion: 75 Resp: 139078
Ion Ratio Lower Upper
75 100
155 78.2 46.9 140.7
157 99.7 60.6 181.7








#93
1,2,4-Trichlorobenzene
Concen: 44.270 ug/l
RT: 15.263 min Scan# 4949
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02


Tgt Ion:180 Resp: 471340
Ion Ratio Lower Upper
180 100
182 95.1 48.3 144.9
145 31.3 14.5 43.5



#94
Hexachlorobutadiene
Concen: 49.735 ug/l
RT: 15.367 min Scan# 4988
Delta R.T. 0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion:225 Resp: 298768
Ion Ratio Lower Upper
225 100
223 63.4 31.9 95.7
227 64.1 32.1 96.3









#95
Naphthalene
Concen: 40.113 ug/l
RT: 15.504 min Scan# 5039
Delta R.T. -0.000 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion:128 Resp: 1177443

Ion	Ratio	Lower	Upper
128	100		
127	13.1	10.5	15.7
129	11.0	8.7	13.1



#96
1,2,3-Trichlorobenzene
Concen: 43.928 ug/l
RT: 15.700 min Scan# 5112
Delta R.T. 0.003 min
Lab File: VN070670.D
Acq: 18 Jan 2022 17:02

Tgt Ion:180 Resp: 447081

Ion	Ratio	Lower	Upper
180	100		
182	96.6	47.4	142.2
145	32.6	15.0	45.1

