

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011822\
 Data File : VN070685.D
 Acq On : 18 Jan 2022 23:39
 Operator : JC/MD
 Sample : N1148-09DL 10X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-49DL

Quant Time: Jan 19 03:42:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1076514	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1722314	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1518487	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	519369	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	707543	45.861	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	91.720%		
35) Dibromofluoromethane	8.021	113	498152	46.628	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	93.260%		
50) Toluene-d8	10.441	98	2133740	47.409	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	94.820%		
62) 4-Bromofluorobenzene	12.731	95	669076	41.718	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	83.440%		
Target Compounds					Qvalue	
11) Tert butyl alcohol	5.372	59	244551	87.859	ug/l	96
16) Acetone	4.293	43	35906	3.353	ug/l	95
19) Methyl tert-butyl Ether	5.618	73	1232597	28.972	ug/l	95
40) Benzene	8.459	78	69661	1.250	ug/l	94
95) Naphthalene	15.501	128	5634	5.137	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011822\
Data File : VN070685.D
Acq On : 18 Jan 2022 23:39
Operator : JC/MD
Sample : N1148-09DL 10X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-49DL

Quant Time: Jan 19 03:42:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

