

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011822\
 Data File : VN070687.D
 Acq On : 19 Jan 2022 00:29
 Operator : JC/MD
 Sample : N1148-06DL 50X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-46DL

Quant Time: Jan 19 03:43:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1136109	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1799714	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1601759	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.672	152	629070	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	737223	45.278	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	90.560%		
35) Dibromofluoromethane	8.021	113	519737	46.556	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	93.120%		
50) Toluene-d8	10.441	98	2239488	47.619	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.240%		
62) 4-Bromofluorobenzene	12.731	95	762271	45.485	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	90.960%		
Target Compounds					Qvalue	
16) Acetone	4.291	43	31502	2.788	ug/l	98
31) Cyclohexane	8.094	56	114844	4.006	ug/l #	83
39) Methylcyclohexane	9.459	83	54310	2.341	ug/l	88
52) Toluene	10.505	92	47997	1.378	ug/l	98
67) Ethyl Benzene	11.843	91	1516877	23.575	ug/l	96
68) m/p-Xylenes	11.950	106	1379908	58.954	ug/l	91
69) o-Xylene	12.278	106	512473	22.062	ug/l	90
73) Isopropylbenzene	12.575	105	89879	1.510	ug/l	100
78) n-propylbenzene	12.919	91	223640	3.360	ug/l	95
80) 1,3,5-Trimethylbenzene	13.058	105	209770	4.359	ug/l	98
84) 1,2,4-Trimethylbenzene	13.367	105	1335910	28.878	ug/l	98
95) Naphthalene	15.507	128	175717	9.802	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011822\
Data File : VN070687.D
Acq On : 19 Jan 2022 00:29
Operator : JC/MD
Sample : N1148-06DL 50X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-46DL

Quant Time: Jan 19 03:43:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

