

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011822\
 Data File : VN070670.D
 Acq On : 18 Jan 2022 17:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 19 01:10:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	0.578	0.525	9.2	96	0.00
3 P	Chloromethane	0.845	0.740	12.4	95	0.00
4 C	Vinyl Chloride	0.803	0.716	10.8#	93	0.00
5 T	Bromomethane	0.418	0.406	2.9	101	0.00
6 T	Chloroethane	0.501	0.447	10.8	96	0.00
7 T	Trichlorofluoromethane	0.918	0.865	5.8	100	0.00
8 T	Diethyl Ether	0.401	0.358	10.7	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.555	0.493	11.2	97	0.00
10 T	Methyl Iodide	0.715	0.662	7.4	93	0.00
11 T	Tert butyl alcohol	0.129	0.116	10.1	95	0.00
12 CM	1,1-Dichloroethene	0.553	0.489	11.6#	97	0.00
13 T	Acrolein	0.185	0.156	15.7	94	0.00
14 T	Allyl chloride	1.205	1.134	5.9	95	0.00
15 T	Acrylonitrile	0.411	0.376	8.5	91	0.00
16 T	Acetone	0.497	0.403	18.9	83	0.00
17 T	Carbon Disulfide	1.502	1.386	7.7	95	0.00
18 T	Methyl Acetate	0.978	0.850	13.1	92	0.00
19 T	Methyl tert-butyl Ether	1.976	1.860	5.9	95	0.00
20 T	Methylene Chloride	0.655	0.588	10.2	95	0.00
21 T	trans-1,2-Dichloroethene	0.557	0.524	5.9	96	0.00
22 T	Diisopropyl ether	2.389	2.250	5.8	95	0.00
23 T	Vinyl Acetate	2.025	1.940	4.2	95	0.00
24 P	1,1-Dichloroethane	1.195	1.107	7.4	95	0.00
25 T	2-Butanone	0.626	0.556	11.2	88	0.00
26 T	2,2-Dichloropropane	0.916	0.863	5.8	97	0.00
27 T	cis-1,2-Dichloroethene	0.661	0.612	7.4	95	0.00
28 T	Bromochloromethane	0.536	0.491	8.4	97	0.00
29 T	Tetrahydrofuran	0.383	0.353	7.8	92	0.00
30 C	Chloroform	1.342	1.070	20.3#	97	0.00
31 T	Cyclohexane	1.262	1.124	10.9	95	0.00
32 T	1,1,1-Trichloroethane	0.927	0.871	6.0	95	0.00
33 S	1,2-Dichloroethane-d4	0.717	0.625	12.8	90	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.310	0.285	8.1	91	0.00
36 T	1,1-Dichloropropene	0.522	0.507	2.9	97	0.00
37 T	Ethyl Acetate	0.693	0.656	5.3	92	0.00
38 T	Carbon Tetrachloride	0.464	0.449	3.2	95	0.00
39 T	Methylcyclohexane	0.645	0.641	0.6	96	0.00
40 TM	Benzene	1.618	1.546	4.4	96	0.00
41 T	Methacrylonitrile	0.353	0.328	7.1	93	0.00
42 TM	1,2-Dichloroethane	0.561	0.539	3.9	96	0.00
43 T	Isopropyl Acetate	1.059	1.018	3.9	94	0.00
44 TM	Trichloroethene	0.367	0.353	3.8	96	0.00
45 C	1,2-Dichloropropane	0.440	0.423	3.9#	96	0.00
46 T	Dibromomethane	0.256	0.247	3.5	95	0.00
47 T	Bromodichloromethane	0.523	0.505	3.4	96	0.00
48 T	Methyl methacrylate	0.504	0.507	-0.6	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011822\
 Data File : VN070670.D
 Acq On : 18 Jan 2022 17:02
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Jan 19 01:10:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	98	0.00
50 S	Toluene-d8	1.307	1.194	8.6	91	0.00
51 T	4-Methyl-2-Pentanone	0.676	0.655	3.1	92	0.00
52 CM	Toluene	0.968	0.948	2.1#	96	0.00
53 T	t-1,3-Dichloropropene	0.561	0.575	-2.5	97	0.00
54 T	cis-1,3-Dichloropropene	0.627	0.630	-0.5	96	0.00
55 T	1,1,2-Trichloroethane	0.380	0.362	4.7	95	0.00
56 T	Ethyl methacrylate	0.639	0.638	0.2	94	0.00
57 T	1,3-Dichloropropane	0.673	0.654	2.8	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.151	0.158	-4.6	95	0.00
59 T	2-Hexanone	0.491	0.484	1.4	90	0.00
60 T	Dibromochloromethane	0.355	0.358	-0.8	96	0.00
61 T	1,2-Dibromoethane	0.382	0.368	3.7	94	0.00
62 S	4-Bromofluorobenzene	0.466	0.443	4.9	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	101	0.00
64 T	Tetrachloroethene	0.367	0.346	5.7	96	0.00
65 PM	Chlorobenzene	1.083	1.036	4.3	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.373	0.360	3.5	95	0.00
67 C	Ethyl Benzene	2.008	1.974	1.7#	96	0.00
68 T	m/p-Xylenes	0.731	0.727	0.5	96	0.00
69 T	o-Xylene	0.725	0.717	1.1	96	0.00
70 T	Styrene	1.160	1.192	-2.8	97	0.00
71 P	Bromoform	0.282	0.280	0.7	95	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	4.730	4.325	8.6	97	0.00
74 T	N-amyl acetate	1.934	1.843	4.7	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.598	1.402	12.3	93	0.00
76 T	1,2,3-Trichloropropane	1.326	1.304	1.7	109	0.00
77 T	Bromobenzene	1.061	0.974	8.2	96	0.00
78 T	n-propylbenzene	5.290	5.113	3.3	98	0.00
79 T	2-Chlorotoluene	3.346	3.085	7.8	97	0.00
80 T	1,3,5-Trimethylbenzene	3.825	3.606	5.7	97	0.00
81 T	trans-1,4-Dichloro-2-butene	0.445	0.421	5.4	95	0.00
82 T	4-Chlorotoluene	3.127	3.017	3.5	98	0.00
83 T	tert-Butylbenzene	3.330	3.101	6.9	96	0.00
84 T	1,2,4-Trimethylbenzene	3.677	3.531	4.0	97	0.00
85 T	sec-Butylbenzene	4.527	4.418	2.4	98	0.00
86 T	p-Isopropyltoluene	3.556	3.558	-0.1	98	0.00
87 T	1,3-Dichlorobenzene	1.764	1.707	3.2	97	0.00
88 T	1,4-Dichlorobenzene	1.756	1.660	5.5	98	0.00
89 T	n-Butylbenzene	2.934	3.001	-2.3	100	0.00
90 T	Hexachloroethane	0.622	0.594	4.5	97	0.00
91 T	1,2-Dichlorobenzene	1.743	1.627	6.7	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.271	0.239	11.8	90	0.00
93 T	1,2,4-Trichlorobenzene	0.774	0.810	-4.7	97	0.00
94 T	Hexachlorobutadiene	0.516	0.514	0.4	101	0.00
95 T	Naphthalene	2.001	2.024	-1.1	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011822\
Data File : VN070670.D
Acq On : 18 Jan 2022 17:02
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Jan 19 01:10:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.738	0.768	-4.1	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6