

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011823\
 Data File : VN076289.D
 Acq On : 18 Jan 2023 11:10
 Operator : JC\MD
 Sample : VN0118MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0118MBL01

Quant Time: Jan 19 00:45:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	329571	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	514926	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	431284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	166856	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	169045	45.810	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.620%
35) Dibromofluoromethane	8.177	113	152841	47.450	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.900%
50) Toluene-d8	10.570	98	578247	48.609	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.220%
62) 4-Bromofluorobenzene	12.853	95	174100	44.831	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011823\
Data File : VN076289.D
Acq On : 18 Jan 2023 11:10
Operator : JC\MD
Sample : VN0118MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0118MBL01

Quant Time: Jan 19 00:45:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

