

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011823\
 Data File : VN076305.D
 Acq On : 18 Jan 2023 19:31
 Operator : JC\MD
 Sample : 01190-09
 Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PE-8

Quant Time: Jan 19 00:54:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	430693	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	659692	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	554435	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	242527	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	216675	44.931	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.860%
35) Dibromofluoromethane	8.171	113	186768	45.259	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.520%
50) Toluene-d8	10.571	98	747574	49.053	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.100%
62) 4-Bromofluorobenzene	12.853	95	237457	47.728	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011823\
Data File : VN076305.D
Acq On : 18 Jan 2023 19:31
Operator : JC\MD
Sample : 01190-09
Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PE-8

Quant Time: Jan 19 00:54:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

