

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011823\
 Data File : VN076305.D
 Acq On : 18 Jan 2023 19:31
 Operator : JC\MD
 Sample : 01190-09
 Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PE-8

Quant Time: Jan 19 00:54:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	430693	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	659692	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	554435	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	242527	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	216675	44.931	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.860%
35) Dibromofluoromethane	8.171	113	186768	45.259	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.520%
50) Toluene-d8	10.571	98	747574	49.053	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.100%
62) 4-Bromofluorobenzene	12.853	95	237457	47.728	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.460%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011823\
Data File : VN076305.D
Acq On : 18 Jan 2023 19:31
Operator : JC\MD
Sample : 01190-09
Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PE-8

Quant Time: Jan 19 00:54:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1049
Delta R.T. -0.006 min
Lab File: VN076305.D
Acq: 18 Jan 2023 19:31




Instrument :
MSVOA_N
ClientSampleId :
PE-8


Tgt Ion:168 Resp: 430693
Ion Ratio Lower Upper
168 100
99 54.7 41.6 62.4



#33
1,2-Dichloroethane-d4
Concen: 44.931 ug/l
RT: 8.583 min Scan# 1110
Delta R.T. 0.001 min
Lab File: VN076305.D
Acq: 18 Jan 2023 19:31

Tgt Ion: 65 Resp: 216675
Ion Ratio Lower Upper
65 100
67 52.5 0.0 104.6





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.100 min Scan# 1199

Delta R.T. -0.006 min

Lab File: VN076305.D

Acq: 18 Jan 2023 19:31

Instrument :

MSVOA_N

ClientSampleId :

PE-8





Tgt Ion:114 Resp: 659692

Ion Ratio Lower Upper

114 100

63 18.4 0.0 35.8

88 14.9 0.0 28.0



#35

Dibromofluoromethane

Concen: 45.259 ug/l

RT: 8.171 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VN076305.D

Acq: 18 Jan 2023 19:31




Tgt Ion:113 Resp: 186768


Ion Ratio Lower Upper

113 100

111 104.3 82.7 124.1

192 20.6 15.6 23.4









#50
Toluene-d8
Concen: 49.053 ug/l
RT: 10.571 min Scan# 1448
Delta R.T. -0.000 min
Lab File: VN076305.D
Acq: 18 Jan 2023 19:31




Instrument :
MSVOA_N
ClientSampleId :
PE-8


Tgt Ion: 98 Resp: 747574
Ion Ratio Lower Upper
98 100
100 65.4 52.2 78.4



#62
4-Bromofluorobenzene
Concen: 47.728 ug/l
RT: 12.853 min Scan# 1836
Delta R.T. 0.006 min
Lab File: VN076305.D
Acq: 18 Jan 2023 19:31

Tgt Ion: 95 Resp: 237457
Ion Ratio Lower Upper
95 100
174 86.5 0.0 174.6
176 82.6 0.0 167.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.871 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN076305.D
Acq: 18 Jan 2023 19:31

Instrument :
MSVOA_N
ClientSampleId :
PE-8

Tgt Ion:117 Resp: 554435
Ion Ratio Lower Upper
117 100
82 53.1 42.2 63.2
119 31.9 26.9 40.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 1996
Delta R.T. -0.000 min
Lab File: VN076305.D
Acq: 18 Jan 2023 19:31

Tgt Ion:152 Resp: 242527
Ion Ratio Lower Upper
152 100
115 58.3 29.6 88.9
150 157.5 0.0 351.4

