

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011921\
 Data File : VN065549.D
 Acq On : 19 Jan 2021 11:29
 Operator : JC/MD
 Sample : VN0119WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0119WBL01

Quant Time: Jan 20 00:27:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 18 18:11:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.627	168	208663	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	337998	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	363281	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	154609	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.994	65	104571	54.26	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.52%	
35) Dibromofluoromethane	7.553	113	104443	51.75	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.50%	
50) Toluene-d8	10.058	98	419998	55.45	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 110.90%	
62) 4-Bromofluorobenzene	12.376	95	160778	56.56	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 113.12%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011921\
Data File : VN065549.D
Acq On : 19 Jan 2021 11:29
Operator : JC/MD
Sample : VN0119WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0119WBL01

Quant Time: Jan 20 00:27:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011821W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 18 18:11:15 2021
Response via : Initial Calibration

