

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011921\
 Data File : VN065554.D
 Acq On : 19 Jan 2021 13:52
 Operator : JC/MD
 Sample : M1159-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BORE-2-ALT

Quant Time: Jan 20 00:28:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 18 18:11:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.630	168	203632	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	334187	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	370010	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	162122	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	104529	55.57	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 111.14%	
35) Dibromofluoromethane	7.556	113	102817	51.53	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 103.06%	
50) Toluene-d8	10.061	98	420258	56.11	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 112.22%	
62) 4-Bromofluorobenzene	12.376	95	165251	58.79	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 117.58%	
Target Compounds						
16) Acetone	3.791	43	11733	17.61	ug/l	90
20) Methylene Chloride	4.508	84	17214	8.39	ug/l	93
43) Isopropyl Acetate	8.135	43	14311	3.75	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011921\
Data File : VN065554.D
Acq On : 19 Jan 2021 13:52
Operator : JC/MD
Sample : M1159-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BORE-2-ALT

Quant Time: Jan 20 00:28:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011821W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 18 18:11:15 2021
Response via : Initial Calibration

