

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011921\
 Data File : VN065555.D
 Acq On : 19 Jan 2021 14:16
 Operator : JC/MD
 Sample : M1169-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-C

Quant Time: Jan 20 00:29:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 18 18:11:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.631	168	207847	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.553	114	339078	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.380	117	374021	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.318	152	155804	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.991	65	106516	55.48	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.96%
35) Dibromofluoromethane	7.553	113	104557	51.65	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.30%
50) Toluene-d8	10.061	98	427517	56.26	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	112.52%
62) 4-Bromofluorobenzene	12.376	95	162412	56.95	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	113.90%
Target Compounds						
					Qvalue	
16) Acetone	3.795	43	8769	12.44	ug/l	96
20) Methylene Chloride	4.508	84	14922	7.13	ug/l	98
43) Isopropyl Acetate	8.135	43	14213	3.67	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011921\
Data File : VN065555.D
Acq On : 19 Jan 2021 14:16
Operator : JC/MD
Sample : M1169-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-C

Quant Time: Jan 20 00:29:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011821W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 18 18:11:15 2021
Response via : Initial Calibration

