

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011921\
 Data File : VN065556.D
 Acq On : 19 Jan 2021 15:21
 Operator : JC/MD
 Sample : M1169-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-C

Quant Time: Jan 20 00:29:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011821W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jan 18 18:11:15 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.624	168	297806	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.550	114	415349	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.376	117	379342	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.315	152	153919	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.988	65	134659	48.95	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.90%
35) Dibromofluoromethane	7.550	113	121946	49.17	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.34%
50) Toluene-d8	10.058	98	492471	52.91	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.82%
62) 4-Bromofluorobenzene	12.373	95	155767	44.59	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.18%
Target Compounds						
					Qvalue	
16) Acetone	3.788	43	8779	8.18	ug/l #	87
20) Methylene Chloride	4.505	84	14541	4.85	ug/l	94
43) Isopropyl Acetate	8.129	43	13826	2.92	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011921\
Data File : VN065556.D
Acq On : 19 Jan 2021 15:21
Operator : JC/MD
Sample : M1169-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-C

Quant Time: Jan 20 00:29:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011821W.M
Quant Title : SW846 8260
QLast Update : Mon Jan 18 18:11:15 2021
Response via : Initial Calibration

