

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011922\
 Data File : VN070692.D
 Acq On : 19 Jan 2022 11:49
 Operator : JC/MD
 Sample : VN0119WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0119WBL01

Quant Time: Jan 20 03:34:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1087104	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1745627	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1519866	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	542817	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	714986	45.892	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.780%
35) Dibromofluoromethane	8.021	113	504099	46.555	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.100%
50) Toluene-d8	10.440	98	2147397	47.075	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.160%
62) 4-Bromofluorobenzene	12.731	95	680005	41.833	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	83.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011922\
Data File : VN070692.D
Acq On : 19 Jan 2022 11:49
Operator : JC/MD
Sample : VN0119WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0119WBL01

Quant Time: Jan 20 03:34:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

