

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011922\
 Data File : VN070698.D
 Acq On : 19 Jan 2022 14:30
 Operator : JC/MD
 Sample : N1178-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 AUD-1244

Quant Time: Jan 20 03:38:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1025433	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1625728	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1417471	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	500347	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	683867	46.534	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.060%
35) Dibromofluoromethane	8.021	113	473950	46.998	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.000%
50) Toluene-d8	10.441	98	2016761	47.472	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	94.940%
62) 4-Bromofluorobenzene	12.731	95	635123	41.954	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	83.900%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.374	59	17553	6.620	ug/l #	79
16) Acetone	4.291	43	162041	15.886	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011922\
Data File : VN070698.D
Acq On : 19 Jan 2022 14:30
Operator : JC/MD
Sample : N1178-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
AUD-1244

Quant Time: Jan 20 03:38:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

