

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011922\
 Data File : VN070699.D
 Acq On : 19 Jan 2022 14:55
 Operator : JC/MD
 Sample : PB142114TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB142114TB

Quant Time: Jan 20 03:38:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1052149	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1681447	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1526572	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	541210	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	700125	46.431	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.860%
35) Dibromofluoromethane	8.021	113	485318	46.531	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.060%
50) Toluene-d8	10.440	98	2109380	48.007	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.020%
62) 4-Bromofluorobenzene	12.731	95	686073	43.818	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.640%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	54092	5.168	ug/l	99
20) Methylene Chloride	5.098	84	26296	1.907	ug/l #	81
43) Isopropyl Acetate	8.560	43	106409	2.989	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011922\
Data File : VN070699.D
Acq On : 19 Jan 2022 14:55
Operator : JC/MD
Sample : PB142114TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB142114TB

Quant Time: Jan 20 03:38:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

