

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011922\
 Data File : VN070706.D
 Acq On : 19 Jan 2022 17:53
 Operator : JC/MD
 Sample : N1183-03
 Misc : 1.05g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1336-1137

Quant Time: Jan 20 03:42:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	1152239	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1869143	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	1493728	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	633347	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	758409	45.927	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.860%
35) Dibromofluoromethane	8.018	113	519989	44.849	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	89.700%
50) Toluene-d8	10.440	98	2211162	45.270	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	90.540%
62) 4-Bromofluorobenzene	12.731	95	708430	40.702	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	81.400%
Target Compounds						
					Qvalue	
16) Acetone	4.315	43	38996	3.402	ug/l	93
18) Methyl Acetate	4.862	43	43760	1.941	ug/l #	87
25) 2-Butanone	7.351	43	118730	8.232	ug/l	92
95) Naphthalene	15.504	128	39514	6.033	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011922\
Data File : VN070706.D
Acq On : 19 Jan 2022 17:53
Operator : JC/MD
Sample : N1183-03
Misc : 1.05g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1336-1137

Quant Time: Jan 20 03:42:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

