

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011923\
 Data File : VN076317.D
 Acq On : 19 Jan 2023 12:55
 Operator : JC\MD
 Sample : 01214-19
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TJRC-011823-GW-DUP

Quant Time: Jan 20 00:54:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	452392	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	711295	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	616824	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	258578	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	239548	47.291	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.580%
35) Dibromofluoromethane	8.171	113	216011	48.548	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.100%
50) Toluene-d8	10.570	98	809599	49.269	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.540%
62) 4-Bromofluorobenzene	12.853	95	259702	48.412	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.820%
Target Compounds						
30) Chloroform	7.976	83	18768	2.137	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011923\
Data File : VN076317.D
Acq On : 19 Jan 2023 12:55
Operator : JC\MD
Sample : 01214-19
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TJRC-011823-GW-DUP

Quant Time: Jan 20 00:54:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

