

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011923\
 Data File : VN076324.D
 Acq On : 19 Jan 2023 15:43
 Operator : JC\MD
 Sample : 01221-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Jan 20 00:57:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	400125	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	629545	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	544110	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	231918	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	210158	46.909	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.820%
35) Dibromofluoromethane	8.171	113	187890	47.711	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.420%
50) Toluene-d8	10.571	98	721659	49.620	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.240%
62) 4-Bromofluorobenzene	12.853	95	224470	47.278	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.560%
Target Compounds						
95) Naphthalene	15.647	128	3574	2.125	ug/l	Qvalue # 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011923\
Data File : VN076324.D
Acq On : 19 Jan 2023 15:43
Operator : JC\MD
Sample : 01221-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Jan 20 00:57:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

