

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011923\
 Data File : VN076331.D
 Acq On : 19 Jan 2023 18:31
 Operator : JC\MD
 Sample : 01236-01 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 NWB-1875

Quant Time: Jan 20 01:00:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	413577	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	648325	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	575081	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	236078	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	217501	46.969	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.940%
35) Dibromofluoromethane	8.177	113	196077	48.348	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.700%
50) Toluene-d8	10.571	98	746260	49.825	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.660%
62) 4-Bromofluorobenzene	12.853	95	237147	48.501	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.000%
Target Compounds						
					Qvalue	
52) Toluene	10.635	92	4515	0.392	ug/l	82
68) m/p-Xylenes	12.071	106	4483	0.537	ug/l	98
84) 1,2,4-Trimethylbenzene	13.482	105	3942	0.257	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011923\
Data File : VN076331.D
Acq On : 19 Jan 2023 18:31
Operator : JC\MD
Sample : 01236-01 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
NWB-1875

Quant Time: Jan 20 01:00:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

