

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011923\
 Data File : VN076333.D
 Acq On : 19 Jan 2023 19:19
 Operator : JC\MD
 Sample : VN0119MBS01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0119MBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/20/2023
 Supervised By : Mahesh Dadoda 01/20/2023

Quant Time: Jan 20 01:02:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	328776	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	536978	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	479486	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	223302	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	183992	49.981	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.960%		
35) Dibromofluoromethane	8.171	113	164357	48.930	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.860%		
50) Toluene-d8	10.571	98	608998	49.092	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	98.180%		
62) 4-Bromofluorobenzene	12.853	95	207950	51.349	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	102.700%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.142	85	67950	20.168	ug/l	98
3) Chloromethane	2.377	50	67036	19.217	ug/l	93
4) Vinyl Chloride	2.524	62	84963	19.864	ug/l	99
5) Bromomethane	2.953	94	79447	22.681	ug/l	96
6) Chloroethane	3.124	64	64241	20.029	ug/l	96
7) Trichlorofluoromethane	3.506	101	116156	19.739	ug/l	94
8) Diethyl Ether	3.977	74	40873	20.708	ug/l	98
9) 1,1,2-Trichlorotrifluo...	4.383	101	63320	19.008	ug/l	99
10) Methyl Iodide	4.600	142	100411	19.733	ug/l	98
11) Tert butyl alcohol	5.542	59	50032	101.610	ug/l	99
12) 1,1-Dichloroethene	4.359	96	61914	19.661	ug/l	96
13) Acrolein	4.195	56	56275	95.567	ug/l	100
14) Allyl chloride	5.036	41	67693	18.446	ug/l	97
15) Acrylonitrile	5.730	53	140102	99.988	ug/l	97
16) Acetone	4.442	43	110742	93.795	ug/l	94
17) Carbon Disulfide	4.724	76	135226	17.246	ug/l	95
18) Methyl Acetate	5.042	43	78068	19.520	ug/l	100
19) Methyl tert-butyl Ether	5.806	73	217194	21.186	ug/l	99
20) Methylene Chloride	5.289	84	76585	21.590	ug/l	97
21) trans-1,2-Dichloroethene	5.794	96	63798	18.701	ug/l	98
22) Diisopropyl ether	6.683	45	176377	20.380	ug/l	97
23) Vinyl Acetate	6.618	43	562077	100.493	ug/l	99
24) 1,1-Dichloroethane	6.577	63	112743	19.544	ug/l	98
25) 2-Butanone	7.494	43	173463	96.242	ug/l	99
26) 2,2-Dichloropropane	7.500	77	87437	17.984	ug/l	97
27) cis-1,2-Dichloroethene	7.494	96	77255	19.416	ug/l	100
28) Bromochloromethane	7.824	49	50352	21.535	ug/l	96
29) Tetrahydrofuran	7.847	42	114079	102.553	ug/l	98
30) Chloroform	7.971	83	128317	20.103	ug/l	100
31) Cyclohexane	8.259	56	103677	18.334	ug/l	97
32) 1,1,1-Trichloroethane	8.177	97	116321	19.960	ug/l	92
36) 1,1-Dichloropropene	8.377	75	88513	18.544	ug/l	98
37) Ethyl Acetate	7.565	43	75814	19.825	ug/l	99
38) Carbon Tetrachloride	8.365	117	101476	19.065	ug/l	98
39) Methylcyclohexane	9.606	83	112019	19.218	ug/l	96
40) Benzene	8.612	78	275801	19.324	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011923\
 Data File : VN076333.D
 Acq On : 19 Jan 2023 19:19
 Operator : JC\MD
 Sample : VN0119MBS01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0119MBS01

Manual Integrations
 APPROVED

Reviewed By : John Carlone 01/20/2023
 Supervised By : Mahesh Dadoda 01/20/2023

Quant Time: Jan 20 01:02:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 17 03:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.788	41	38821	19.366	ug/l	95
42) 1,2-Dichloroethane	8.677	62	94047	19.259	ug/l	99
43) Isopropyl Acetate	8.694	43	124449	19.982	ug/l	99
44) Trichloroethene	9.359	130	78272	19.390	ug/l	90
45) 1,2-Dichloropropane	9.630	63	65931	19.294	ug/l	97
46) Dibromomethane	9.712	93	51266	19.871	ug/l	98
47) Bromodichloromethane	9.888	83	99415	19.868	ug/l	98
48) Methyl methacrylate	9.682	41	58168	19.728	ug/l	97
49) 1,4-Dioxane	9.700	88	26645	428.272	ug/l #	96
51) 4-Methyl-2-Pentanone	10.447	43	375698	102.758	ug/l	99
52) Toluene	10.635	92	184356	19.316	ug/l	99
53) t-1,3-Dichloropropene	10.841	75	95149	19.947	ug/l	99
54) cis-1,3-Dichloropropene	10.318	75	107481	19.585	ug/l	96
55) 1,1,2-Trichloroethane	11.018	97	73340	20.103	ug/l	96
56) Ethyl methacrylate	10.877	69	101215	21.490	ug/l	98
57) 1,3-Dichloropropane	11.165	76	120115	20.260	ug/l	98
58) 2-Chloroethyl Vinyl ether	10.165	63	138483	87.095	ug/l	98
59) 2-Hexanone	11.200	43	271709	102.553	ug/l	100
60) Dibromochloromethane	11.359	129	78963	20.561	ug/l	99
61) 1,2-Dibromoethane	11.471	107	72676	19.958	ug/l	98
64) Tetrachloroethene	11.106	164	83049	19.206	ug/l	98
65) Chlorobenzene	11.894	112	195533	19.291	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.959	131	73888	19.593	ug/l	98
67) Ethyl Benzene	11.965	91	347172	20.013	ug/l	99
68) m/p-Xylenes	12.076	106	279661	40.205	ug/l	98
69) o-Xylene	12.400	106	137816	20.175	ug/l	98
70) Styrene	12.412	104	219383	20.114	ug/l	99
71) Bromoform	12.582	173	55773	20.479	ug/l #	99
73) Isopropylbenzene	12.700	105	353166	20.444	ug/l	100
74) N-amyl acetate	12.494	43	105012	21.133	ug/l	98
75) 1,1,2,2-Tetrachloroethane	12.941	83	98775	19.890	ug/l	98
76) 1,2,3-Trichloropropane	12.994	75	90095m	21.609	ug/l	
77) Bromobenzene	12.982	156	85534	20.851	ug/l	100
78) n-propylbenzene	13.035	91	385356	19.770	ug/l	100
79) 2-Chlorotoluene	13.129	91	237906	19.890	ug/l	100
80) 1,3,5-Trimethylbenzene	13.176	105	299169	20.672	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.741	75	24083	17.944	ug/l	95
82) 4-Chlorotoluene	13.129	91	237906	19.890	ug/l	100
83) tert-Butylbenzene	13.441	119	275562	21.409	ug/l	97
84) 1,2,4-Trimethylbenzene	13.482	105	297409	20.495	ug/l	100
85) sec-Butylbenzene	13.618	105	360374	20.177	ug/l	100
86) p-Isopropyltoluene	13.729	119	303233	20.647	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	153051	18.798	ug/l	99
88) 1,4-Dichlorobenzene	13.812	146	151385	19.607	ug/l	100
89) n-Butylbenzene	14.059	91	222358	19.072	ug/l	98
90) Hexachloroethane	14.341	117	49829	19.658	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	156715	19.325	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.723	75	16492	19.105	ug/l	97
93) 1,2,4-Trichlorobenzene	15.394	180	71625	19.107	ug/l	99
94) Hexachlorobutadiene	15.506	225	39654	18.426	ug/l	98
95) Naphthalene	15.647	128	212892	19.513	ug/l	100
96) 1,2,3-Trichlorobenzene	15.847	180	68011	18.942	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011923\
Data File : VN076333.D
Acq On : 19 Jan 2023 19:19
Operator : JC\MD
Sample : VN0119MBSD01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0119MBSD01

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023

Quant Time: Jan 20 01:02:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN011923\
Data File : VN076333.D
Acq On : 19 Jan 2023 19:19
Operator : JC\MD
Sample : VN0119MBS01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 25 Sample Multiplier: 1


Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01

Manual Integrations
APPROVED

Reviewed By : John Carlone 01/20/2023
Supervised By : Mahesh Dadoda 01/20/2023


Quant Time: Jan 20 01:02:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011623W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 17 03:59:22 2023
Response via : Initial Calibration





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 10
Delta R.T. -0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01



Tgt Ion:168 Resp: 328776
Ion Ratio Lower Upper
168 100
99 55.3 41.6 62.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023



#2
Dichlorodifluoromethane
Concen: 20.168 ug/l
RT: 2.142 min Scan# 15
Delta R.T. 0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Tgt Ion: 85 Resp: 67950
Ion Ratio Lower Upper
85 100
87 35.0 17.1 51.2









#3
Chloromethane
Concen: 19.217 ug/l
RT: 2.377 min Scan# 51
Delta R.T. 0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Instrument :
MSVOA_N
ClientSampleId :
VN0119MBSD01

Tgt Ion: 50 Resp: 67030
Ion Ratio Lower Upper
50 100
52 34.6 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023



#4
Vinyl Chloride
Concen: 19.864 ug/l
RT: 2.524 min Scan# 80
Delta R.T. -0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19


Tgt Ion: 62 Resp: 84963
Ion Ratio Lower Upper
62 100
64 32.6 25.8 38.6





#5
Bromomethane
Concen: 22.681 ug/l
RT: 2.953 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01




Tgt Ion: 94 Resp: 7944
Ion Ratio Lower Upper
94 100
96 97.9 75.1 112.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023




#6
Chloroethane
Concen: 20.029 ug/l
RT: 3.124 min Scan# 182
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Tgt Ion: 64 Resp: 64241
Ion Ratio Lower Upper
64 100
66 33.4 24.8 37.2





#7
 Trichlorofluoromethane
 Concen: 19.739 ug/l
 RT: 3.506 min Scan# 248
 Delta R.T. -0.006 min
 Lab File: VN076333.D
 Acq: 19 Jan 2023 19:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0119MBSD01




Tgt Ion:101 Resp: 116150
 Ion Ratio Lower Upper
 101 100
 103 67.8 50.2 75.4

Manual Integrations
 APPROVED



Reviewed By :John Carlone 01/20/2023
 Supervised By :Mahesh Dadoda 01/20/2023




#8
 Diethyl Ether
 Concen: 20.708 ug/l
 RT: 3.977 min Scan# 327
 Delta R.T. 0.006 min
 Lab File: VN076333.D
 Acq: 19 Jan 2023 19:19



Tgt Ion: 74 Resp: 40873
 Ion Ratio Lower Upper
 74 100
 45 77.1 37.8 113.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.008 ug/l
RT: 4.383 min Scan# 397
Delta R.T. -0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Instrument :
MSVOA_N
ClientSampleId :
VN0119MBSD01

Tgt Ion:101 Resp: 63320
Ion Ratio Lower Upper
101 100
85 44.7 35.5 53.3
151 82.7 65.5 98.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023



#10
Methyl Iodide
Concen: 19.733 ug/l
RT: 4.600 min Scan# 433
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Tgt Ion:142 Resp: 100411
Ion Ratio Lower Upper
142 100
127 44.6 35.0 52.4
141 13.6 11.5 17.3









#11
Tert butyl alcohol
Concen: 101.610 ug/l
RT: 5.542 min Scan# 591
Delta R.T. 0.012 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01

Tgt Ion: 59 Resp: 5003
Ion Ratio Lower Upper
59 100
57 10.4 8.5 12.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023



#12
1,1-Dichloroethene
Concen: 19.661 ug/l
RT: 4.359 min Scan# 392
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Tgt Ion: 96 Resp: 61914
Ion Ratio Lower Upper
96 100
61 132.4 108.2 162.4
98 63.4 47.0 70.6





#13

Acrolein

Concen: 95.567 ug/l

RT: 4.195 min Scan# 30

Delta R.T. 0.006 min

Lab File: VN076333.D

Acq: 19 Jan 2023 19:19

Instrument :

MSVOA_N

ClientSampleId :

VN0119MBS01

Tgt Ion: 56 Resp: 5627

Ion Ratio Lower Upper

56 100





55 70.0 55.8 83.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/20/2023

Supervised By :Mahesh Dadoda 01/20/2023



#14

Allyl chloride

Concen: 18.446 ug/l

RT: 5.036 min Scan# 507

Delta R.T. -0.000 min

Lab File: VN076333.D

Acq: 19 Jan 2023 19:19




Tgt Ion: 41 Resp: 67693


Ion Ratio Lower Upper

41 100

39 87.5 67.4 101.0


76 50.4 39.8 59.8





#15
Acrylonitrile
Concen: 99.988 ug/l
RT: 5.730 min Scan# 61
Delta R.T. -0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01




Tgt Ion: 53 Resp: 14010
Ion Ratio Lower Upper
53 100
52 83.3 64.7 97.1
51 34.6 29.2 43.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023




#16
Acetone
Concen: 93.795 ug/l
RT: 4.442 min Scan# 406
Delta R.T. -0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Tgt Ion: 43 Resp: 110742
Ion Ratio Lower Upper
43 100
58 36.9 26.7 40.1





#17
Carbon Disulfide
Concen: 17.246 ug/l
RT: 4.724 min Scan# 41
Delta R.T. -0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01




Tgt Ion: 76 Resp: 135220
Ion Ratio Lower Upper
76 100
78 8.1 8.0 12.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023




#18
Methyl Acetate
Concen: 19.520 ug/l
RT: 5.042 min Scan# 508
Delta R.T. 0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19



Tgt Ion: 43 Resp: 78068
Ion Ratio Lower Upper
43 100
74 29.2 23.3 34.9









#19
Methyl tert-butyl Ether
Concen: 21.186 ug/l
RT: 5.806 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01

Tgt Ion: 73 Resp: 217194
Ion Ratio Lower Upper
73 100
57 20.3 15.9 23.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023



#20
Methylene Chloride
Concen: 21.590 ug/l
RT: 5.289 min Scan# 550
Delta R.T. -0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19


Tgt Ion: 84 Resp: 76585
Ion Ratio Lower Upper
84 100
49 101.6 84.5 126.7
51 33.6 26.0 39.0
86 67.5 52.0 78.0





#21
trans-1,2-Dichloroethene
Concen: 18.701 ug/l
RT: 5.794 min Scan# 636
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01



Tgt Ion: 96 Resp: 63798
Ion Ratio Lower Upper
96 100
61 120.0 97.0 145.4
98 62.9 51.9 77.9




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023



#22
Diisopropyl ether
Concen: 20.380 ug/l
RT: 6.683 min Scan# 787
Delta R.T. -0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Tgt Ion: 45 Resp: 176377
Ion Ratio Lower Upper
45 100
43 52.6 44.4 66.6
87 36.0 28.4 42.6
59 14.4 11.1 16.7









#23
 Vinyl Acetate
 Concen: 100.493 ug/l
 RT: 6.618 min Scan# 775
 Delta R.T. 0.006 min
 Lab File: VN076333.D
 Acq: 19 Jan 2023 19:19

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0119MBS01

Tgt Ion: 43 Resp: 56207
 Ion Ratio Lower Upper
 43 100
 86 13.7 11.2 16.8




Manual Integrations
 APPROVED


Reviewed By :John Carlone 01/20/2023
 Supervised By :Mahesh Dadoda 01/20/2023



#24
 1,1-Dichloroethane
 Concen: 19.544 ug/l
 RT: 6.577 min Scan# 769
 Delta R.T. -0.006 min
 Lab File: VN076333.D
 Acq: 19 Jan 2023 19:19


Tgt Ion: 63 Resp: 112743
 Ion Ratio Lower Upper
 63 100
 98 9.1 3.9 11.7
 100 4.3 2.2 6.6





#25
2-Butanone
Concen: 96.242 ug/l
RT: 7.494 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01




Tgt Ion: 43 Resp: 17346
Ion Ratio Lower Upper
43 100
72 33.9 27.4 41.2

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023




#26
2,2-Dichloropropane
Concen: 17.984 ug/l
RT: 7.500 min Scan# 926
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19



Tgt Ion: 77 Resp: 87437
Ion Ratio Lower Upper
77 100
97 24.8 11.8 35.4





#27

cis-1,2-Dichloroethene

Concen: 19.416 ug/l

RT: 7.494 min Scan# 91

Delta R.T. 0.000 min

Lab File: VN076333.D

Acq: 19 Jan 2023 19:19

Instrument :

MSVOA_N

ClientSampleId :

VN0119MBS01

Tgt Ion: 96 Resp: 7725

Ion Ratio Lower Upper

96 100

61 129.7 0.0 259.6





98 64.5 0.0 129.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/20/2023

Supervised By :Mahesh Dadoda 01/20/2023



#28

Bromochloromethane

Concen: 21.535 ug/l

RT: 7.824 min Scan# 981

Delta R.T. -0.000 min

Lab File: VN076333.D

Acq: 19 Jan 2023 19:19




Tgt Ion: 49 Resp: 50352


Ion Ratio Lower Upper

49 100

129 2.5 0.0 5.2


130 90.9 76.0 114.0





#29
Tetrahydrofuran
Concen: 102.553 ug/l
RT: 7.847 min Scan# 985
Delta R.T. 0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Instrument :
MSVOA_N
ClientSampleId :
VN0119MBSD01




Tgt Ion: 42 Resp: 114079
Ion Ratio Lower Upper
42 100
72 57.0 44.0 66.0
71 51.2 41.1 61.7

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023




#30
Chloroform
Concen: 20.103 ug/l
RT: 7.971 min Scan# 1006
Delta R.T. -0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Tgt Ion: 83 Resp: 128317
Ion Ratio Lower Upper
83 100
85 64.1 51.0 76.6





#31
Cyclohexane
Concen: 18.334 ug/l
RT: 8.259 min Scan# 1055
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01




Tgt Ion: 56 Resp: 10367
Ion Ratio Lower Upper
56 100
69 31.1 25.8 38.8
84 95.8 79.0 118.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023




#32
1,1,1-Trichloroethane
Concen: 19.960 ug/l
RT: 8.177 min Scan# 1041
Delta R.T. 0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19



Tgt Ion: 97 Resp: 116321
Ion Ratio Lower Upper
97 100
99 60.8 55.0 82.4
61 42.0 36.2 54.4





#33

1,2-Dichloroethane-d4

Concen: 49.981 ug/l

RT: 8.583 min Scan# 1110

Delta R.T. 0.001 min

Lab File: VN076333.D

Acq: 19 Jan 2023 19:19

Instrument :

MSVOA_N

ClientSampleId :

VN0119MBS01

Tgt Ion: 65 Resp: 183992

Ion Ratio Lower Upper

65 100





67 52.7 0.0 104.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/20/2023

Supervised By :Mahesh Dadoda 01/20/2023



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1199

Delta R.T. 0.000 min

Lab File: VN076333.D

Acq: 19 Jan 2023 19:19




Tgt Ion:114 Resp: 536978


Ion Ratio Lower Upper

114 100

63 17.6 0.0 35.8


88 14.7 0.0 28.0





#35
Dibromofluoromethane
Concen: 48.930 ug/l
RT: 8.171 min Scan# 1040
Delta R.T. -0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Instrument :
MSVOA_N
ClientSampleId :
VN0119MBSD01




Tgt Ion:113 Resp: 16435
Ion Ratio Lower Upper
113 100
111 101.7 82.7 124.1
192 20.0 15.6 23.4

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023




#36
1,1-Dichloropropene
Concen: 18.544 ug/l
RT: 8.377 min Scan# 1075
Delta R.T. -0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19



Tgt Ion: 75 Resp: 88513
Ion Ratio Lower Upper
75 100
110 39.1 18.8 56.3
77 30.2 25.0 37.4





#37

Ethyl Acetate

Concen: 19.825 ug/l

RT: 7.565 min Scan# 91

Delta R.T. -0.000 min

Lab File: VN076333.D

Acq: 19 Jan 2023 19:19

Instrument :

MSVOA_N

ClientSampleId :

VN0119MBSD01

Tgt Ion: 43 Resp: 75814





Ion	Ratio	Lower	Upper
43	100		
61	14.1	11.1	16.7
70	11.9	10.0	15.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/20/2023

Supervised By :Mahesh Dadoda 01/20/2023



#38

Carbon Tetrachloride

Concen: 19.065 ug/l

RT: 8.365 min Scan# 1073




Delta R.T. -0.000 min


Lab File: VN076333.D

Acq: 19 Jan 2023 19:19

Tgt Ion:117 Resp: 101476

Ion	Ratio	Lower	Upper
117	100		
119	92.7	75.4	113.0
121	28.6	24.2	36.2





#39

Methylcyclohexane

Concen: 19.218 ug/l

RT: 9.606 min Scan# 112019

Delta R.T. 0.006 min

Lab File: VN076333.D

Acq: 19 Jan 2023 19:19

Instrument :

MSVOA_N

ClientSampleId :

VN0119MBS01

Tgt Ion: 83 Resp: 112019

Ion Ratio Lower Upper

83 100

55 60.1 51.7 77.5





98 48.4 40.0 60.0

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/20/2023

Supervised By :Mahesh Dadoda 01/20/2023



#40

Benzene

Concen: 19.324 ug/l

RT: 8.612 min Scan# 1115

Delta R.T. -0.000 min

Lab File: VN076333.D




Acq: 19 Jan 2023 19:19


Tgt Ion: 78 Resp: 275801

Ion Ratio Lower Upper

78 100

77 23.6 18.6 28.0





#41
Methacrylonitrile
Concen: 19.366 ug/l
RT: 7.788 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19





Instrument :
MSVOA_N
ClientSampleId :
VN0119MBSD01

Tgt Ion: 41 Resp: 3882

Ion	Ratio	Lower	Upper
41	100		
39	58.3	42.4	63.6
67	91.2	68.8	103.2
52	35.6	27.7	41.5

Manual Integrations
APPROVED




Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023




#42
1,2-Dichloroethane
Concen: 19.259 ug/l
RT: 8.677 min Scan# 1126
Delta R.T. -0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Tgt Ion: 62 Resp: 94047

Ion	Ratio	Lower	Upper
62	100		
98	10.8	0.0	20.4





#43

Isopropyl Acetate

Concen: 19.982 ug/l

RT: 8.694 min Scan# 1129

Delta R.T. 0.000 min

Lab File: VN076333.D

Acq: 19 Jan 2023 19:19

Instrument :

MSVOA_N

ClientSampleId :

VN0119MBS01

Tgt Ion: 43 Resp: 124449





Ion	Ratio	Lower	Upper
43	100		
61	29.7	23.5	35.3
87	16.4	13.8	20.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/20/2023

Supervised By :Mahesh Dadoda 01/20/2023



#44

Trichloroethene

Concen: 19.390 ug/l

RT: 9.359 min Scan# 1242




Delta R.T. 0.006 min


Lab File: VN076333.D

Acq: 19 Jan 2023 19:19

Tgt Ion:130 Resp: 78272

Ion	Ratio	Lower	Upper
130	100		
95	86.4	0.0	192.0





#45

1,2-Dichloropropane

Concen: 19.294 ug/l

RT: 9.630 min Scan# 1288

Delta R.T. 0.006 min

Lab File: VN076333.D

Acq: 19 Jan 2023 19:19

Instrument :

MSVOA_N

ClientSampleId :

VN0119MBS01

Tgt Ion: 63 Resp: 6593

Ion Ratio Lower Upper

63 100





65 31.4 26.4 39.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/20/2023

Supervised By :Mahesh Dadoda 01/20/2023



#46

Dibromomethane

Concen: 19.871 ug/l

RT: 9.712 min Scan# 1302

Delta R.T. -0.000 min

Lab File: VN076333.D

Acq: 19 Jan 2023 19:19




Tgt Ion: 93 Resp: 51266


Ion Ratio Lower Upper

93 100

95 81.2 66.0 99.0


174 100.0 81.3 121.9





#47
Bromodichloromethane
Concen: 19.868 ug/l
RT: 9.888 min Scan# 1332
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01



Tgt Ion: 83 Resp: 9941
Ion Ratio Lower Upper
83 100
85 66.9 52.1 78.1
127 9.9 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023



#48
Methyl methacrylate
Concen: 19.728 ug/l
RT: 9.682 min Scan# 1297
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Tgt Ion: 41 Resp: 58168
Ion Ratio Lower Upper
41 100
69 95.8 80.8 121.2
39 68.3 54.4 81.6





#49

1,4-Dioxane

Concen: 428.272 ug/l

RT: 9.700 min Scan# 1300

Delta R.T. 0.000 min

Lab File: VN076333.D

Acq: 19 Jan 2023 19:19

Instrument :

MSVOA_N

ClientSampleId :

VN0119MBS01

Tgt Ion: 88 Resp: 2664

Ion Ratio Lower Upper

88 100

43 27.5 18.1 27.1





58 59.6 46.5 69.7

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/20/2023

Supervised By :Mahesh Dadoda 01/20/2023



#50

Toluene-d8

Concen: 49.092 ug/l

RT: 10.571 min Scan# 1448

Delta R.T. -0.000 min

Lab File: VN076333.D




Acq: 19 Jan 2023 19:19


Tgt Ion: 98 Resp: 608998

Ion Ratio Lower Upper

98 100


100 65.2 52.2 78.4





#51
4-Methyl-2-Pentanone
Concen: 102.758 ug/l
RT: 10.447 min Scan# 1427
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01



Tgt Ion: 43 Resp: 375698
Ion Ratio Lower Upper
43 100
58 43.0 35.1 52.7




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023



#52
Toluene
Concen: 19.316 ug/l
RT: 10.635 min Scan# 1459
Delta R.T. 0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Tgt Ion: 92 Resp: 184356
Ion Ratio Lower Upper
92 100
91 169.9 136.6 205.0









#53
t-1,3-Dichloropropene
Concen: 19.947 ug/l
RT: 10.841 min Scan# 1493
Delta R.T. 0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01

Tgt Ion: 75 Resp: 95149
Ion Ratio Lower Upper
75 100
77 32.5 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023



#54
cis-1,3-Dichloropropene
Concen: 19.585 ug/l
RT: 10.318 min Scan# 1405
Delta R.T. 0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19


Tgt Ion: 75 Resp: 107481
Ion Ratio Lower Upper
75 100
77 30.3 26.5 39.7
39 44.5 37.1 55.7





#55
1,1,2-Trichloroethane
Concen: 20.103 ug/l
RT: 11.018 min Scan# 1524
Delta R.T. -0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Instrument :
MSVOA_N
ClientSampleId :
VN0119MBSD01




Tgt Ion: 97 Resp: 73340
Ion Ratio Lower Upper
97 100
83 80.1 67.8 101.6
85 56.6 42.3 63.5
99 64.9 50.6 76.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023




#56
Ethyl methacrylate
Concen: 21.490 ug/l
RT: 10.877 min Scan# 1500
Delta R.T. 0.001 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19



Tgt Ion: 69 Resp: 101215
Ion Ratio Lower Upper
69 100
41 55.1 45.3 67.9
39 35.8 29.3 43.9





#57

1,3-Dichloropropane

Concen: 20.260 ug/l

RT: 11.165 min Scan# 11

Delta R.T. -0.000 min

Lab File: VN076333.D

Acq: 19 Jan 2023 19:19

Instrument :

MSVOA_N

ClientSampleId :

VN0119MBS01

Tgt Ion: 76 Resp: 12011

Ion Ratio Lower Upper

76 100





78 31.7 26.4 39.6

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/20/2023

Supervised By :Mahesh Dadoda 01/20/2023



#58

2-Chloroethyl Vinyl ether

Concen: 87.095 ug/l

RT: 10.165 min Scan# 1379

Delta R.T. -0.000 min

Lab File: VN076333.D




Acq: 19 Jan 2023 19:19


Tgt Ion: 63 Resp: 138483

Ion Ratio Lower Upper

63 100

106 33.9 26.2 39.2





#59

2-Hexanone

Concen: 102.553 ug/l

RT: 11.200 min Scan# 1555

Delta R.T. 0.006 min

Lab File: VN076333.D

Acq: 19 Jan 2023 19:19

Instrument :

MSVOA_N

ClientSampleId :

VN0119MBS01

Tgt Ion: 43 Resp: 271709

Ion Ratio Lower Upper

43 100





58 60.0 30.1 90.3

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/20/2023

Supervised By :Mahesh Dadoda 01/20/2023



#60

Dibromochloromethane

Concen: 20.561 ug/l

RT: 11.359 min Scan# 1582

Delta R.T. -0.000 min

Lab File: VN076333.D




Acq: 19 Jan 2023 19:19


Tgt Ion:129 Resp: 78963

Ion Ratio Lower Upper

129 100


127 78.5 38.7 116.1





#61
1,2-Dibromoethane
Concen: 19.958 ug/l
RT: 11.471 min Scan# 1601
Delta R.T. 0.001 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Instrument :
MSVOA_N
ClientSampleId :
VN0119MBSD01




Tgt Ion:107 Resp: 72670
Ion Ratio Lower Upper
107 100
109 96.6 76.0 114.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023




#62
4-Bromofluorobenzene
Concen: 51.349 ug/l
RT: 12.853 min Scan# 1836
Delta R.T. 0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Tgt Ion: 95 Resp: 207950
Ion Ratio Lower Upper
95 100
174 84.9 0.0 174.6
176 83.2 0.0 167.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.865 min Scan# 1668
Delta R.T. -0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01




Tgt Ion:117 Resp: 479480
Ion Ratio Lower Upper
117 100
82 52.1 42.2 63.2
119 32.1 26.9 40.3

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023




#64
Tetrachloroethene
Concen: 19.206 ug/l
RT: 11.106 min Scan# 1539
Delta R.T. -0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Tgt Ion:164 Resp: 83049
Ion Ratio Lower Upper
164 100
166 130.8 104.0 156.0
129 90.3 76.6 114.8
131 89.7 72.6 109.0





#65
Chlorobenzene
Concen: 19.291 ug/l
RT: 11.894 min Scan# 1673
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01




Tgt Ion:112 Resp: 19553
Ion Ratio Lower Upper
112 100
114 31.4 25.3 37.9

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023




#66
1,1,1,2-Tetrachloroethane
Concen: 19.593 ug/l
RT: 11.959 min Scan# 1684
Delta R.T. -0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19



Tgt Ion:131 Resp: 73888
Ion Ratio Lower Upper
131 100
133 95.8 47.4 142.2
119 67.3 32.3 96.8









#67
Ethyl Benzene
Concen: 20.013 ug/l
RT: 11.965 min Scan# 1104
Delta R.T. -0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Instrument :
MSVOA_N
ClientSampleId :
VN0119MBSD01

Tgt Ion: 91 Resp: 347172
Ion Ratio Lower Upper
91 100
106 31.2 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023



#68
m/p-Xylenes
Concen: 40.205 ug/l
RT: 12.076 min Scan# 1704
Delta R.T. 0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Tgt Ion:106 Resp: 279661
Ion Ratio Lower Upper
106 100
91 193.7 157.5 236.3









#69
o-Xylene
Concen: 20.175 ug/l
RT: 12.400 min Scan# 1759
Delta R.T. -0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01

Tgt Ion:106 Resp: 137810
Ion Ratio Lower Upper
106 100
91 209.0 103.2 309.4




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023



#70
Styrene
Concen: 20.114 ug/l
RT: 12.412 min Scan# 1761
Delta R.T. -0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19


Tgt Ion:104 Resp: 219383
Ion Ratio Lower Upper
104 100
78 47.2 37.4 56.0
103 55.9 44.2 66.2





#71
Bromoform
Concen: 20.479 ug/l
RT: 12.582 min Scan# 1790
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01




Tgt Ion:173 Resp: 5577
Ion Ratio Lower Upper
173 100
175 47.7 24.1 72.4
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 1996
Delta R.T. -0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19



Tgt Ion:152 Resp: 223302
Ion Ratio Lower Upper
152 100
115 58.4 29.6 88.9
150 164.5 0.0 351.4









#73
Isopropylbenzene
Concen: 20.444 ug/l
RT: 12.700 min Scan# 1809
Delta R.T. 0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01

Tgt Ion:105 Resp: 35316
Ion Ratio Lower Upper
105 100
120 26.9 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023



#74
N-ethyl acetate
Concen: 21.133 ug/l
RT: 12.494 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Tgt Ion: 43 Resp: 105012
Ion Ratio Lower Upper
43 100
70 49.4 41.2 61.8
55 26.0 21.0 31.6
61 26.2 21.6 32.4









#75
1,1,2,2-Tetrachloroethane
Concen: 19.890 ug/l
RT: 12.941 min Scan# 1851
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Instrument :
MSVOA_N
ClientSampleId :
VN0119MBSD01

Tgt Ion: 83 Resp: 98775
Ion Ratio Lower Upper
83 100
131 11.3 5.5 16.4
85 65.0 31.7 95.1




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023



#76
1,2,3-Trichloropropane
Concen: 21.609 ug/l m
RT: 12.994 min Scan# 1860
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Tgt Ion: 75 Resp: 90095
Ion Ratio Lower Upper
75 100
77 177.4 75.2 225.6









#77
Bromobenzene
Concen: 20.851 ug/l
RT: 12.982 min Scan# 1858
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01

Tgt Ion:156 Resp: 85534
Ion Ratio Lower Upper
156 100
77 186.9 93.4 280.2
158 95.9 48.4 145.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023



#78
n-propylbenzene
Concen: 19.770 ug/l
RT: 13.035 min Scan# 1867
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Tgt Ion: 91 Resp: 385356
Ion Ratio Lower Upper
91 100
120 23.5 11.9 35.5









#79
2-Chlorotoluene
Concen: 19.890 ug/l
RT: 13.129 min Scan# 1883
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01

Tgt Ion: 91 Resp: 237900
Ion Ratio Lower Upper
91 100
126 34.8 17.4 52.2




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023



#80
1,3,5-Trimethylbenzene
Concen: 20.672 ug/l
RT: 13.176 min Scan# 1891
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Tgt Ion:105 Resp: 299169
Ion Ratio Lower Upper
105 100
120 49.2 25.0 75.0









#81
trans-1,4-Dichloro-2-butene
Concen: 17.944 ug/l
RT: 12.741 min Scan# 1816
Delta R.T. 0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01

Tgt Ion: 75 Resp: 2408
Ion Ratio Lower Upper
75 100
53 101.6 75.8 113.8
89 46.3 37.0 55.6




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023



#82
4-Chlorotoluene
Concen: 19.890 ug/l
RT: 13.129 min Scan# 1883
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19


Tgt Ion: 91 Resp: 237906
Ion Ratio Lower Upper
91 100
126 34.8 17.4 52.2





#83
tert-Butylbenzene
Concen: 21.409 ug/l
RT: 13.441 min Scan# 1936
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01




Tgt Ion:119 Resp: 275562
Ion Ratio Lower Upper
119 100
91 62.5 32.2 96.6
134 25.7 13.6 40.8

Manual Integrations
APPROVED



Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023




#84
1,2,4-Trimethylbenzene
Concen: 20.495 ug/l
RT: 13.482 min Scan# 1943
Delta R.T. 0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Tgt Ion:105 Resp: 297409
Ion Ratio Lower Upper
105 100
120 45.6 22.9 68.5





#85
 sec-Butylbenzene
 Concen: 20.177 ug/l
 RT: 13.618 min Scan# 1966
 Delta R.T. 0.000 min
 Lab File: VN076333.D
 Acq: 19 Jan 2023 19:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0119MBS01




Tgt Ion:105 Resp: 360374
 Ion Ratio Lower Upper
 105 100
 134 20.3 10.3 30.8

Manual Integrations
 APPROVED



Reviewed By :John Carlone 01/20/2023
 Supervised By :Mahesh Dadoda 01/20/2023




#86
 p-Isopropyltoluene
 Concen: 20.647 ug/l
 RT: 13.729 min Scan# 1985
 Delta R.T. 0.000 min
 Lab File: VN076333.D
 Acq: 19 Jan 2023 19:19




Tgt Ion:119 Resp: 303233
 Ion Ratio Lower Upper
 119 100
 134 26.3 13.2 39.6
 91 22.9 11.8 35.3





#87
 1,3-Dichlorobenzene
 Concen: 18.798 ug/l
 RT: 13.735 min Scan# 1986
 Delta R.T. 0.000 min
 Lab File: VN076333.D
 Acq: 19 Jan 2023 19:19




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0119MBS01




Tgt Ion:146 Resp: 15305
 Ion Ratio Lower Upper
 146 100
 111 39.3 19.9 59.7
 148 64.4 31.6 94.8

Manual Integrations
 APPROVED



Reviewed By :John Carlone 01/20/2023
 Supervised By :Mahesh Dadoda 01/20/2023




#88
 1,4-Dichlorobenzene
 Concen: 19.607 ug/l
 RT: 13.812 min Scan# 1999
 Delta R.T. -0.000 min
 Lab File: VN076333.D
 Acq: 19 Jan 2023 19:19



Tgt Ion:146 Resp: 151385
 Ion Ratio Lower Upper
 146 100
 111 39.0 19.1 57.5
 148 63.8 31.9 95.9









#89
n-Butylbenzene
Concen: 19.072 ug/l
RT: 14.059 min Scan# 2041
Delta R.T. -0.000 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Instrument :
MSVOA_N
ClientSampleId :
VN0119MBSD01

Tgt Ion: 91 Resp: 222358
Ion Ratio Lower Upper
91 100
92 54.2 26.3 78.9
134 27.1 13.2 39.5




Manual Integrations
APPROVED


Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023



#90
Hexachloroethane
Concen: 19.658 ug/l
RT: 14.341 min Scan# 2089
Delta R.T. 0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19

Tgt Ion:117 Resp: 49829
Ion Ratio Lower Upper
117 100
201 75.4 37.3 111.9





#91

1,2-Dichlorobenzene

Concen: 19.325 ug/l

RT: 14.106 min Scan# 2049

Delta R.T. -0.000 min

Lab File: VN076333.D

Acq: 19 Jan 2023 19:19

Instrument :

MSVOA_N

ClientSampleId :

VN0119MBS01

Tgt Ion:146 Resp: 15671





Ion	Ratio	Lower	Upper
146	100		
111	40.2	20.3	60.8
148	62.9	31.8	95.4

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/20/2023

Supervised By :Mahesh Dadoda 01/20/2023



#92

1,2-Dibromo-3-Chloropropane

Concen: 19.105 ug/l

RT: 14.723 min Scan# 2154




Delta R.T. 0.000 min


Lab File: VN076333.D

Acq: 19 Jan 2023 19:19

Tgt Ion: 75 Resp: 16492

Ion	Ratio	Lower	Upper
75	100		
155	93.2	45.0	135.0
157	119.4	58.3	174.9





#93

1,2,4-Trichlorobenzene

Concen: 19.107 ug/l

RT: 15.394 min Scan# 2268

Delta R.T. -0.000 min

Lab File: VN076333.D


Acq: 19 Jan 2023 19:19

Instrument :

MSVOA_N

ClientSampleId :

VN0119MBS01



Tgt Ion:180 Resp: 71629




Ion	Ratio	Lower	Upper
180	100		
182	94.6	47.0	141.2
145	30.9	15.2	45.5

Manual Integrations

APPROVED

Reviewed By :John Carlone 01/20/2023

Supervised By :Mahesh Dadoda 01/20/2023



#94

Hexachlorobutadiene

Concen: 18.426 ug/l

RT: 15.506 min Scan# 2287




Delta R.T. -0.000 min


Lab File: VN076333.D

Acq: 19 Jan 2023 19:19

Tgt Ion:225 Resp: 39654


Ion	Ratio	Lower	Upper
225	100		
223	61.2	30.8	92.3
227	65.4	30.9	92.8





#95
Naphthalene
Concen: 19.513 ug/l
RT: 15.647 min Scan# 21289
Delta R.T. 0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19




Instrument :
MSVOA_N
ClientSampleId :
VN0119MBS01




Tgt Ion:128 Resp: 21289
Ion Ratio Lower Upper
128 100
127 12.8 10.3 15.5
129 10.7 8.6 13.0

Manual Integrations
APPROVED

Reviewed By :John Carlone 01/20/2023
Supervised By :Mahesh Dadoda 01/20/2023



#96
1,2,3-Trichlorobenzene
Concen: 18.942 ug/l
RT: 15.847 min Scan# 2345
Delta R.T. 0.006 min
Lab File: VN076333.D
Acq: 19 Jan 2023 19:19



Tgt Ion:180 Resp: 68011
Ion Ratio Lower Upper
180 100
182 96.6 47.5 142.5
145 33.4 15.9 47.6

