

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012021\
 Data File : VN065561.D
 Acq On : 20 Jan 2021 10:48
 Operator : JC/MD
 Sample : VN0120WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0120WBL01

Quant Time: Jan 21 01:20:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N011321W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Jan 14 04:12:08 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.155	128	34474	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.550	114	172403	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.379	117	151365	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.987	65	61346	28.48	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	94.93%
60) 4-Bromofluorobenzene	12.376	95	54809	23.91	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	79.70%
63) Toluene-d8	10.058	98	200863	30.32	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	101.07%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012021\
Data File : VN065561.D
Acq On : 20 Jan 2021 10:48
Operator : JC/MD
Sample : VN0120WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0120WBL01

Quant Time: Jan 21 01:20:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N011321W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Jan 14 04:12:08 2021
Response via : Initial Calibration

