

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012022\
 Data File : VN070711.D
 Acq On : 20 Jan 2022 10:36
 Operator : JC/MD
 Sample : VN0120MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0120MBL01

Quant Time: Jan 21 02:10:28 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	1054989	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	1686588	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1475901	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	525544	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	699446	46.261	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.520%
35) Dibromofluoromethane	8.021	113	491008	46.933	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	93.860%
50) Toluene-d8	10.443	98	2067503	46.910	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	93.820%
62) 4-Bromofluorobenzene	12.731	95	662039	42.154	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	84.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012022\
Data File : VN070711.D
Acq On : 20 Jan 2022 10:36
Operator : JC/MD
Sample : VN0120MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0120MBL01

Quant Time: Jan 21 02:10:28 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

