

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012022\
 Data File : VN070714.D
 Acq On : 20 Jan 2022 12:39
 Operator : JC/MD
 Sample : N1158-02
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 R-1A

Quant Time: Jan 21 02:11:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	1064919	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	1708425	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.744	117	1385441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	577868	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	714137	46.792	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.580%
35) Dibromofluoromethane	8.018	113	486648	45.922	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.840%
50) Toluene-d8	10.440	98	2049284	45.903	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	91.800%
62) 4-Bromofluorobenzene	12.733	95	661851	41.603	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	83.200%
Target Compounds						
25) 2-Butanone	7.353	43	114939	8.623	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012022\
Data File : VN070714.D
Acq On : 20 Jan 2022 12:39
Operator : JC/MD
Sample : N1158-02
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
R-1A

Quant Time: Jan 21 02:11:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

