

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012022\
 Data File : VN070715.D
 Acq On : 20 Jan 2022 13:05
 Operator : JC/MD
 Sample : N1182-01
 Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OR-2-011822

Quant Time: Jan 21 02:12:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jan 18 13:32:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	1051045	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1702895	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	1378069	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	577066	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	697527	46.307	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	92.620%
35) Dibromofluoromethane	8.016	113	471193	44.608	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	89.220%
50) Toluene-d8	10.441	98	2056452	46.213	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	92.420%
62) 4-Bromofluorobenzene	12.731	95	665259	41.953	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	83.900%
Target Compounds						
25) 2-Butanone	7.351	43	117447	8.927	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN012022\
Data File : VN070715.D
Acq On : 20 Jan 2022 13:05
Operator : JC/MD
Sample : N1182-01
Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OR-2-011822

Quant Time: Jan 21 02:12:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N011822W.M
Quant Title : SW846 8260
QLast Update : Tue Jan 18 13:32:07 2022
Response via : Initial Calibration

